

UM-26001-F

DT7837 File I/O Programming Manual

Trademark and Copyright Information

Measurement Computing Corporation, InstaCal, Universal Library, and the Measurement Computing logo are either trademarks or registered trademarks of Measurement Computing Corporation. Refer to the Copyrights & Trademarks section on <u>mccdaq.com/legal</u> for more information about Measurement Computing trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies.

© 2015 Measurement Computing Corporation. All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form by any means, electronic, mechanical, by photocopying, recording, or otherwise without the prior written permission of Measurement Computing Corporation.

Notice

Measurement Computing Corporation does not authorize any Measurement Computing Corporation product for use in life support systems and/or devices without prior written consent from Measurement Computing Corporation. Life support devices/systems are devices or systems that, a) are intended for surgical implantation into the body, or b) support or sustain life and whose failure to perform can be reasonably expected to result in injury. Measurement Computing Corporation products are not designed with the components required, and are not subject to the testing required to ensure a level of reliability suitable for the treatment and diagnosis of people.

Table of Contents

About this Manual
Intended Audience
What You Should Learn from this Manual7
Conventions Used in this Manual7
Related Documents
Where to Get Help 8
Chapter 1: Overview
Introduction
Installing the Software 11
Examples
Summary of Supported File I/O Operations
Chapter 2: Using the File I/O Commands 21
Opening and Closing a File
Analog Input and Input Stream Operations
Opening the Subsystem and Input Stream
Configuring the Input Channels
Configuring the Channel Mask for the Input Stream
Configuring the Sample Clock
Configuring the Trigger that Starts Acquisition
Submitting I/O Requests
Arming and Starting Continuous Operations
Getting the Status of Acquisition
Processing I/O Requests
Dealing with Input Buffers
Enabling Buffer Error Reporting
Stopping Continuous Operations
Cleaning up Resources
Analog Output and Output Stream Operations
Performing a Single Value Operation
Opening the Subsystem
Updating the Value of the Analog Output Channel
Closing the Subsystem
Performing a Continuous Output Operation
Opening the Output Stream
Configuring the Channel Mask for the Output Stream
Configuring the Sample Clock
Configuring the Trigger that Starts the Output Operation

Submitting I/O Requests
Arming and Starting Continuous Operations
Getting the Status of the Output Operation
Processing I/O Requests
Dealing with Output Buffers
Enabling Buffer Error Reporting 35
Stopping Continuous Operations
Cleaning up Resources
Tachometer Operations 39
Opening the Subsystem
Configuring the Tachometer Subsystem
Counter/Timer Operations
Opening the Subsystem
Configuring the Counter/Timer Subsystem
Mode
Event Counting Mode
Rate Generation Mode
Non-Retriggerable One-Shot Mode
Idle Mode
Gate
C/T Clock Input Sources
Pulse Output Period, Pulse Width, and Polarity
Starting the Counter/Timer Operation47
Reading the Counter/Timer 47
Stopping the Counter/Timer Operation
Closing the Subsystem
Measure Counter Operations
Opening the Subsystem
Configuring the Measure Counter Subsystem
Digital Input Operations
Opening the Subsystem
Reading the Value
Closing the Subsystem
Digital Output Operations
Opening the Subsystem
Updating the Value of the Port
Closing the Subsystem
Calibration
Analog Input Calibration
Opening the Subsystem55
Calibrating the Offset55

	56
Closing the Subsystem	56
Analog Output Calibration	56
Opening the Subsystem	57
Calibrating the Offset	57
Calibrating the Gain	58
Closing the Subsystem	58
Medifeine the Clete of the Harry EDe on the Medule	
Modifying the State of the User LEDs on the Module	
Sending Data to or Receiving Data from the Host USB Application	61
Opening the File	61
Sending Data to the USB Host	61
Receiving Data from the USB Host	62
Eliminating Data from the Endpoint	62
Closing the File	62
Ŭ	
Chapter 3: File I/O Command Reference	63
close	65
ioctl - IOCTL_ACQ_STATUS_GET	66
ioctl - IOCTL_AIN_CNF_GET	67
ioctl - IOCTL_AIN_CNF_SET	69
ioctl - IOCTL_ARM_SUBSYS	71
ioctl - IOCTL_CHAN_MASK_GET	72
ioctl - IOCTL_CHAN_MASK_SET	
ioctl - IOCTL_CT_CFG_GET	78
	~ -
loctl - IOCTL_CT_CFG_SET	85
ioctl - IOCTL_CI_CFG_SET	85 92
ioctl - IOCTL_CT_CFG_SET	85 92 94
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_SET ioctl - IOCTL_GAIN_POT_WIPER_GET	85 92 94 97
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_SET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_FD_CET	85 92 94 97 99
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_SET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET	85 92 94 97 97 99 101
ioctl - IOCTL_CT_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET	85 92 94 97 99 101 103 105
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_SET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET	85 92 94 97 97 99 101 103 105 11
ioctl - IOCTL_CT_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_SET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_MCTR_CFG_SET	85 92 94 97 99 101 103 105 111 117
ioctl - IOCTL_CT_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET	85 92 94 97 97 99 101 103 105 111 117 119
ioctl - IOCTL_CT_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_SET	85 92 94 97 99 101 103 105 111 117 119 121
ioctl - IOCTL_CIT_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET	85 92 94 97 99 101 103 105 111 117 119 121 123
ioctl - IOCTL_CT_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_OFFSET_POT_WIPER_SET ioctl - IOCTL_OFFSET_POT_WIPER_SET ioctl - IOCTL_OFFSET_POT_WIPER_SET	85 92 94 97 99 101 103 105 111 117 119 121 123 125
ioctl - IOCTL_CT_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_OFFSET_POT_WIPER_SET ioctl - IOCTL_SAMPLE_CLK_GET ioctl - IOCTL_SAMPLE_CLK_SET	85 92 94 97 99 101 103 105 111 117 117 119 121 123 125 127
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_SAMPLE_CLK_GET ioctl - IOCTL_START_SUBSYS	85 92 94 97 99 101 103 105 111 117 119 121 125 125 127 129
ioctl - IOCTL_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_SET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_OFFSET_POT_WIPER_SET ioctl - IOCTL_SAMPLE_CLK_GET ioctl - IOCTL_START_SUBSYS ioctl - IOCTL_START_TRIG_CNF_GET	85 92 94 97 99 101 103 105 111 117 117 121 123 125 127 129 131
ioctl - IOCTL_CT_CFG_SET. ioctl - IOCTL_GAIN_POT_GET. ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET. ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET. ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET. ioctl - IOCTL_OFFSET_POT_GET. ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_OFFSET_POT_WIPER_SET. ioctl - IOCTL_SAMPLE_CLK_GET ioctl - IOCTL_START_SUBSYS. ioctl - IOCTL_START_TRIG_CNF_GET	85 92 94 97 99 101 103 105 111 117 119 121 123 125 127 129 131 135
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_OFFSET_POT_WIPER_SET ioctl - IOCTL_SAMPLE_CLK_SET ioctl - IOCTL_START_SUBSYS ioctl - IOCTL_START_TRIG_CNF_SET ioctl - IOCTL_START_TRIG_CNF_SET	85 92 94 97 99 101 103 105 111 117 117 121 123 125 127 129 131 135 139
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_MCTR_CFG_SET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_OFFSET_POT_WIPER_SET ioctl - IOCTL_SAMPLE_CLK_GET ioctl - IOCTL_START_TRIG_CNF_GET ioctl - IOCTL_START_TRIG_CNF_SET ioctl - IOCTL_START_TRIG_CNF_SET	85 92 94 97 99 101 103 105 111 117 117 123 125 127 129 131 135 139 140
ioctl - IOCTL_CI_CFG_SET ioctl - IOCTL_GAIN_POT_GET ioctl - IOCTL_GAIN_POT_WIPER_GET ioctl - IOCTL_GAIN_POT_WIPER_SET ioctl - IOCTL_LED_GET ioctl - IOCTL_LED_SET ioctl - IOCTL_MCTR_CFG_GET ioctl - IOCTL_OFFSET_POT_GET ioctl - IOCTL_OFFSET_POT_SET ioctl - IOCTL_OFFSET_POT_WIPER_GET ioctl - IOCTL_OFFSET_POT_WIPER_SET. ioctl - IOCTL_OFFSET_POT_WIPER_SET ioctl - IOCTL_SAMPLE_CLK_GET ioctl - IOCTL_START_TRIG_CNF_GET ioctl - IOCTL_START_TRIG_CNF_SET ioctl - IOCTL_START_CFG_GET ioctl - IOCTL_TACH_CFG_GET ioctl - IOCTL_TACH_CFG_SET.	85 92 94 97 99 101 103 105 111 117 119 121 123 125 127 129 131 135 139 140 142

Chapter 4: Programming Flowcharts Using the File I/O Commands
Input Stream Asynchronous Read Operations
Analog Output Synchronous Write Operation151
Output Stream Asynchronous Write Operations
Digital Input Synchronous Read Operation154
Digital Output Synchronous Write Operation
Counter/Timer Operation
Calibration
User LED Modifications
Sending Data to a USB Host
Receiving Data from a USB Host
Chapter 5: Product Support
Index

About this Manual

This manual describes the file I/O commands that are supported by the DT7837 module in Linux user space.

Intended Audience

This document is intended for experienced Linux programmers.

What You Should Learn from this Manual

This manual provides detailed information about the file I/O operations that are supported for the DT7837 module. This manual is organized as follows:

- Chapter 1, "Overview," provides an overview of the file I/O support for the DT7837 module.
- Chapter 2, "Using the File I/O Commands," describes the I/O operations that are supported by the DT98337 module and the file I/O commands that are used to perform these operations.
- Chapter 3, "File I/O Command Reference," provides in-depth information about each of the file I/O commands that are supported by the DT7837 module, including the command syntax, arguments, and examples.
- Chapter 4, "Programming Flowcharts Using the File I/O Commands," provides a flowchart showing the order in which to use the file I/O commands to perform the supported I/O functions of the DT7837 module.
- Chapter 5, "Product Support," provides information about obtaining technical support.
- An index completes this manual.

Conventions Used in this Manual

The following conventions are used in this manual:

- Notes provide useful information or information that requires special emphasis, cautions provide information to help you avoid losing data or damaging your equipment, and warnings provide information to help you avoid catastrophic damage to yourself or your equipment.
- Command syntax is shown in courier font.

Related Documents

Refer to the following documents for more information:

- DT7837 Getting Started help file on our website (hhttp://www.mccdaq.com/PDFs/Manuals/DT7837_WebHelp/ DT7837_Getting_Started.htm)
- *DT7837 User's Manual* (UM-26000). This manual describes the operation of the DT7837 module, as well calibration instructions, operating specifications, and pin assignments.
- Linux documentation on the AIO model (such as *//code.google.com/p/kernel/wiki/AIOUserGuide*) and on Linux file I/O commands.

Where to Get Help

Should you run into problems installing or using the DT7837 module, the Data Translation Technical Support Department is available to provide technical assistance. Refer to Chapter 5 for more information. If you are outside the United States or Canada, call your local distributor, whose number is listed on Data Translation's web site (www.mccdaq.com).



Overview

Introduction	10
Installing the Software	11
Examples	12

Introduction

The Linux operating system provides a powerful and flexible infrastructure for developing embedded software for the DT7837. Because it is supported by an open source community, developers can decrease development time by leveraging source code that is freely available, and reduce costs by taking advantage of software that does not require licensing or distribution fees.

Figure 1 shows how the software support is organized for the DT7837 module.



Figure 1: DT7837 Software Support

The device drivers (DT78xx driver and DT78xx USB driver), which are loaded when the module is powered up, expose the functionality of the module to Linux user space applications using virtual file interfaces, such as **/dev/dt7837-ain** for the analog input subsystem and**/dev/dt7837-ep1out** for a USB gadget OUT endpoint.

This manual describes the virtual files and the file I/O commands that are supported by the DT78xx device driver and DT78xx USB gadget driver. Using these files and file I/O commands, you can write applications in Linux user space that communicate with the subsystems on the DT7837 module to perform I/O operations.

Installing the Software

Refer to the *DT7837 Getting Started* help file on the Data Translation website for detailed information on installing and setting up a host system for use in developing application programs for the DT7837 module.

Examples

Once you have installed the DT7837 software to your Linux host computer, you can access the DT7837 example programs, which are located in the following subdirectory under the TI SDK directory: **example-applications/dt78xx-examples/**

The example programs demonstrate the capabilities of the hardware and illustrate how user programs communicate with the DT7837 drivers using file I/O commands and IOCTLS. All examples are written in ANSI C and are open source.

Each example program is located in a separate subdirectory. Each example directory includes a README.txt file that provides information about the example and a makefile, which allows you to build the example using the TI SDK toolchain and cross compiler.

You can build these example programs, use them to test the hardware, and modify them as needed to get up and running quickly. It is recommended that you refer to these examples when learning about the file I/O commands.



Figure 2: DT7837 Example Programs

The example programs are summarized in Table 1.

Board Features	Example Directory	Description	
Analog Input/ Input Stream	aio-in	Performs an asynchronous analog input operation and stores the data to a file.	
	data-logger	Acquires data from all the channels in the input stream and stores it to a file. Many file formats are supported.	
	fir-filter	Performs an input stream and an output stream operation simultaneously and continuously. Data from the input stream is filtered and then output from the analog output channel.	
	sig-analyzer	An embedded web server and signal analyzer. It acquires data from two analog input channels and performs and FFT on the data. The results are displayed to a client's web browser when connected to the device.	
	web-server	An embedded web server that performs most of the functions of the DT7837 module and saves the acquired data to a file.	
Analog output/ Output stream	aout-single	Performs a synchronous write operation on the analog output channel.	
	fir-filter	Performs an input stream and an output stream operation simultaneously and continuously. Data from the input stream is filtered and then output from the analog output channel.	
	function-gen	Performs an asynchronous analog output operation, generating one of the following waveforms using Direct Digital Synthesis: sine, triangle, sawtooth, or square. This example also outputs a pulse waveform on several bits of the digital output port.	
Analog Input and Analog Output	dt7837-calibration	A command-line program that calibrates the analog input and analog output circuitry of the DT7837 module.	
Calibration	web-calibration	A web-based calibration program that calibrates the analog input and analog output circuitry of the DT7837 module.	
Digital I/O	digio	Performs a synchronous write operation on the digital output port, and then reads back the value that was output by performing a synchronous read operation on the digital input port.	
	web-server	An embedded web server that performs most of the functions of the DT7837 module and saves the acquired data to a file.	
	function-gen	Performs an asynchronous analog output operation, generating one of the following waveforms using Direct Digital Synthesis: sine, triangle, sawtooth, or square. This example also outputs a pulse waveform on several bits of the digital output port.	

Table 1: Description of the Examples

Board Features	Example Directory	Description	
Counter/timer	clk-gen	Uses the counter/timer to generate an output clock.	
event-counter Performs an event counting operation.		Performs an event counting operation.	
	web-server	An embedded web server that performs most of the functions of the DT7837 module and saves the acquired data to a file.	
USB	usb-loopback	Demonstrates use of the gadget USB driver to send data from a host application to the DT7837 using and OUT USB pipe and to receive data from the DT7837 using IN USB pipe.	
Misc	common	Includes common functions used by the other example programs.	

Table 1: Description of the Examples (cont.)

Note: Linux is an open-source development environment. As such, our example programs may use code from other vendors. This code is for demonstration purposes only. If you want to use this code for commercial purposes, you must ensure that you resolve any licensing issues with the appropriate parties.

Refer to the *DT7837 Getting Started* help file on the Data Translation website for more information about building these example programs.

Summary of Supported File I/O Operations

Table 2 lists the supported file I/O operations for the DT7837 module.

Subsystem	Virtual File	File I/O Command	Description
Analog Input /dev/d	/dev/dt7837-ain	open	Opens the subsystem.
		close	Closes the subsystem.
		ioctl - IOCTL_AIN_CFG_SET	Configures the parameters of the analog input subsystem.
		ioctl - IOCTL_AIN_CFG_GET	Returns the configuration of the analog input subsystem.
		ioctl - IOCTL_GAIN_POT_SET	Sets the gain potentiometer for a specified analog input channel and calibration register.
		ioctl - IOCTL_GAIN_POT_GET	Returns the configuration of the gain potentiometer for a specified analog input channel and calibration register.
		ioctl - IOCTL_GAIN_POT_WIPER_SET	Sets the wiper value for the gain potentiometer associated with a specified analog input channel.
		ioctl - IOCTL_GAIN_POT_WIPER_GET	Returns the configuration of the wiper value for the gain potentiometer associated with a specified analog input channel.
		ioctl - IOCTL_OFFSET_POT_SET	Sets the offset potentiometer for a specified analog input channel and calibration register.
		ioctl - IOCTL_OFFSET_POT_GET	Returns the configuration of the offset potentiometer for a specified analog input channel and calibration register.
		ioctl - IOCTL_OFFSET_POT_WIPER_SET	Sets the wiper value for the offset potentiometer associated with a specified analog input channel.
		ioctl - IOCTL_OFFSET_POT_WIPER_GET	Returns the configuration of the wiper value for the offset potentiometer associated with a specified analog input channel.
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.
			ioctl - IOCTL_LED_GET

Subsystem	Virtual File	File I/O Command	Description					
Input Stream	/dev/dt7837-stream-in	open	Opens the input stream.					
		close	Closes the input stream.					
		ioctl - IOCTL_SAMPLE_CLK_SET	Configures the sample clock frequency for the input stream.					
	ioctl - IOCTL_SAMPLE_CLK_GET	Returns the configuration of the sample clock frequency for the input stream.						
		ioctl - IOCTL_START_TRIG_CFG_SET	Configures the start trigger for the input stream.					
		ioctl - IOCTL_START_TRIG_CFG_GET	Returns the configured start trigger for the input stream.					
		ioctl - IOCTL_CHAN_MASK_SET	Configures the channels from which to acquire data.					
	ioctl - IOCTL_CHAN_MASK_GET	Returns the channels that are enabled for acquisition.						
	ioctl - IOCTL_ARM_SUBSYS	Arms the module to detect a trigger for the input stream.						
	ioctl - IOCTL_START_SUBSYS	Starts continuous acquisition (an asynchronous operation) when a software trigger is specified.						
	ioctl - IOCTL_ACQ_STATUS_GET	Returns the status of acquisition (armed or triggered).						
		ioctl - IOCTL_STOP_SUBSYS	Stops continuous acquisition.					
		io_setup	Opens an asynchronous I/O context.					
		io_submit	Submits I/O requests.					
	io_getevents io_cancel io_destroy	io_getevents	io_getevents	Processes events when I/O requests are completed.				
		io_cancel	Cancels I/O requests.					
								io_destroy
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.					
		ioctl - IOCTL_LED_GET	Returns the state of the user LEDs on the module.					

Subsystem	Virtual File	File I/O Command	Description
Analog Output	/dev/dt7837-aout	open	Opens the subsystem.
		close	Closes the subsystem.
		ioctl - IOCTL_GAIN_POT_SET	Sets the gain potentiometer for a specified analog output channel and calibration register.
		ioctl - IOCTL_GAIN_POT_GET	Returns the configuration of the gain potentiometer for a specified analog output channel and calibration register.
		ioctl - IOCTL_GAIN_POT_WIPER_SET	Sets the wiper value for the gain potentiometer associated with a specified analog output channel.
		ioctl - IOCTL_GAIN_POT_WIPER_GET	Returns the configuration of the wiper value for the gain potentiometer associated with a specified analog output channel.
		ioctl - IOCTL_OFFSET_POT_SET	Sets the offset potentiometer for a specified analog output channel and calibration register.
		ioctl - IOCTL_OFFSET_POT_GET	Returns the configuration of the offset potentiometer for a specified analog output channel and calibration register.
		ioctl - IOCTL_OFFSET_POT_WIPER_SET	Sets the wiper value for the offset potentiometer associated with a specified analog output channel.
		ioctl - IOCTL_OFFSET_POT_WIPER_GET	Returns the configuration of the wiper value for the offset potentiometer associated with a specified analog output channel.
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.
			ioctl - IOCTL_LED_GET

Subsystem	Virtual File	File I/O Command	Description												
Output	/dev/dt7837-stream-out	open	Opens the input stream.												
Stream		close	Closes the input stream.												
		ioctl - IOCTL_SAMPLE_CLK_SET	Configures the sample clock frequency for the output stream.												
		ioctl - IOCTL_SAMPLE_CLK_GET	Returns the configuration of the sample clock frequency for the output stream.												
		ioctl - IOCTL_START_TRIG_CFG_SET	Configures the start trigger for the output stream.												
		ioctl - IOCTL_START_TRIG_CFG_GET	Returns the configured start trigger for the output stream.												
		ioctl - IOCTL_CHAN_MASK_SET	Configures the channels that you want to update in the output stream.												
		ioctl - IOCTL_CHAN_MASK_GET	Returns the output channels that are enabled in the analog output stream.												
		ioctl - IOCTL_ARM_SUBSYS	Arms the module to detect a trigger for the output stream.												
		ioctl - IOCTL_START_SUBSYS	Starts a continuous output operation (an asynchronous operation) when a software trigger is specified.												
														ioctl - IOCTL_ACQ_STATUS_GET	Returns the status of the output operation (armed or triggered).
		ioctl - IOCTL_STOP_SUBSYS	Stops the output operation.												
	/dev/dt7837-din	io_setup	Opens an asynchronous I/O context.												
		io_submit	Submits I/O requests.												
				io_getevents	io_getevents	Processes events when I/O requests are completed.									
		io_cancel	Cancels I/O requests.												
			io_destroy	Destroys an asynchronous I/O context.											
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.												
		ioctl - IOCTL_LED_GET	Returns the state of the user LEDs on the module.												
Digital Input		open	Opens the subsystem.												
		close	Closes the subsystem.												
		read	Synchronously reads a single value from the digital input port.												

Subsystem	Virtual File	File I/O Command	Description
Digital Input (cont.)	/dev/dt7837-din	ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.
		ioctl - IOCTL_LED_GET	Returns the state of the user LEDs on the module.
Digital Output	/dev/dt7837-dout	open	Opens the subsystem.
		close	Closes the subsystem.
		write	Synchronously writes a single value to the digital output port.
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.
		ioctl - IOCTL_LED_GET	Returns the state of the user LEDs on the module.
Counter/	/dev/dt7837-ctr-tmr	open	Opens the subsystem.
Timer		close	Closes the subsystem.
		ioclt - IOCTL_CT_CFG_SET	Configures the parameters of the counter/timer subsystem.
		ioclt - IOCTL_CT_CFG_GET	Returns the configuration of the counter/timer subsystem.
		ioctl - IOCTL_START_SUBSYS	Starts the counter/timer operation.
		ioctl - IOCTL_STOP_SUBSYS	Stops the counter/timer operation.
		read	Synchronously reads a single value from the counter/timer.
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.
		ioctl - IOCTL_LED_GET	Returns the state of the user LEDs on the module.
Measure	/dev/dt7837-measure	open	Opens the subsystem.
Counter		close	Closes the subsystem.
		ioctl - IOCTL_MCTR_CFG_SET	Configures the parameters of the measure counter subsystem.
		ioctl - IOCTL_MCTR_CFG_GET	Returns the configuration of the measure counter subsystem.
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.
		ioctl - IOCTL_LED_GET	Returns the state of the user LEDs on the module.

Subsystem	Virtual File	File I/O Command	Description
Tachometer	/dev/dt7837-tach	open	Opens the subsystem.
		close	Closes the subsystem.
		ioctl - IOCTL_TACH_CFG_SET	Configures the parameters of the tachometer subsystem.
		ioctl - IOCTL_TACH_CFG_GET	Returns the configuration of the tachometer subsystem.
		ioctl - IOCTL_LED_SET	Turns the user LEDs on the module either on or off.
		ioctl - IOCTL_LED_GET	Returns the state of the user LEDs on the module.
Gadget USB IN endpoints	/dev/dt7837-ep1in /dev/dt7837-ep2in	open	Opens the IN file.
		close	Closes the IN file.
		write	Performs a blocking write operation to send data to a USB host.
		aio_write	Performs a non-blocking, asynchronous write operation to send data to a USB host.
		flush	Removes the data from the IN endpoint.
Gadget USB OUT endpoints	/dev/dt7837-ep1out /dev/dt7837-ep2out /dev/dt7837-ep3out /dev/dt7837-ep4out /dev/dt7837-ep5out	open	Opens the OUT file.
		close	Closes the OUT file.
		read	Performs a blocking read operation to receive data from a USB host.
		aio_read	Performs a non-blocking, asynchronous read operation to receive data from a USB host.
		flush	Removes the data from the IN endpoint.



Using the File I/O Commands

Opening and Closing a File.	. 22
Analog Input and Input Stream Operations	. 24
Analog Output and Output Stream Operations	. 32
Fachometer Operations	. 39
Counter/Timer Operations	. 41
Measure Counter Operations	. 48
Digital Input Operations	. 52
Digital Output Operations	. 53
Calibration	. 54
Modifying the State of the User LEDs on the Module	. 59
Sending Data to or Receiving Data from the Host USB Application	. 61

Opening and Closing a File

The DT7837 device driver currently supports the following files for performing embedded I/O operations on the module:

- /dev/dt7837-ain Analog input subsystem. You can configure the parameters of each analog input channel of the analog input subsystem and then read the continuous values from the analog input subsystem using the input stream file (/dev/DT7837-stream-in).
- /dev/dt7837-aout Analog output subsystem. You can configure the parameters of each analog output channel of the analog output subsystem and then perform a synchronous write operation to output a single value from the analog output subsystem or output continuous values from the analog output subsystem using the output stream file (/dev/DT7837-stream-out).
- /dev/dt7837-din Digital input subsystem. This subsystem supports synchronous read operations where a single value is read from the digital input port. To read the digital input data with the analog input data in the input stream, use the /dev/DT7837-stream-in file instead.
- /dev/dt7837-dout Digital output subsystem. This subsystem supports synchronous write operations where a single value is written to the digital output port.
- /dev/dt7837-tach Tachometer input subsystem. You can configure the parameters of the tachometer input channel. You can then read the value of the tachometer using a synchronous read operation, or you can read the tachometer input with the analog input data in the input stream using the /dev/DT7837-stream-in file instead.
- /dev/dt7837-measure Measure counter subsystem. You can configure the parameters of the measure counter. You can then read the value of the measure counter using a synchronous read operation, or you can read the measure counter with the analog input data in the input stream using the /dev/DT7837-stream-in file instead.
- /dev/dt7837-ctr-tmr Counter/timer subsystem. You can configure the parameters of the measure counter. You can then read the value of the measure counter using a synchronous read operation, or you can read the measure counter with the analog input data in the input stream using the /dev/DT7837-stream-in file instead.
- /dev/dt7837-stream-in Input stream. Use this file to perform a continuous input operation from each input channel specified in the input stream. Channels can include the analog inputs, digital input port, tachometer input, measure counter, and counter/timer.
- /dev/dt7837-stream-out Output stream. Use this file to perform a continuous output operation from each input channel specified in the output stream. Channels can include the analog output channels and the digital output port.

The DT78xx USB gadget driver supports these virtual files for sending data to and receiving data from the USB host application:

- /dev/dt7837-ep1in EP1 IN endpoint, address 0x81.
- /dev/dt7837-ep1out EP1 OUT endpoint, address 0x01.
- /dev/dt7837-ep2in EP2 IN endpoint, address 0x82.
- /dev/dt7837-ep2out EP2 OUT endpoint, address 0x02.
- /dev/dt7837-ep3out EP3 OUT endpoint, address 0x03.

- /dev/dt7837-ep4out EP4 OUT endpoint, address 0x04.
- /dev/dt7837-ep5out EP5 OUT endpoint, address 0x05.

Before you can perform an I/O operation, you must open the file that you want to read from or write to using the Linux file I/O command **open**.

You specify the pathname of the file to open and a flag that determines how to open the file, and the command returns a file descriptor that is used in subsequent calls to the file. The following flags determine how the file is opened:

- O_RDONLY –Read only
- O_WRONLY Write only
- O_RDWR Read and write

When you finished performing the I/O operation, you must close each subsystem and/or stream that was opened using the Linux file I/O command **close**.

Analog Input and Input Stream Operations

The DT7837 module supports continuous analog input operations using the input stream.

You can also acquire data from other input channels besides the analog input channels in the input stream, including the digital input port, tachometer, and measure counter. This section describes analog input and input stream operations on the DT7837 module.

Opening the Subsystem and Input Stream

To perform an analog input operation, open the file corresponding to the analog input subsystem (/dev/dt7837-ain) and in the input stream (/dev/dt7837-stream-in) using the Linux command **open**, described on page 22.

Configuring the Input Channels

To configure the analog input channels of the analog input subsystem, use the **ioctl - IOCTL_AIN_CFG_SET** command.

You specify the file descriptor for the analog input subsystem and then specify the parameters for each channel of the subsystem. Parameters include the following:

- Analog input channel to configure (ain) Specify analog input channel 0, 1, 2, or 3.
- Gain (*gain*) Specify either a gain of 1 for an effective input range of ±10 V or a gain of 10 for an effective input range of ±1 V.
- Coupling type (*ac_coupling*) Specify either DC coupling or AC coupling for the analog input channel.
- Current (current_on) Specify either that the 4 mA current source is on or that it is off.
- Input type (*differential*) Specify either the single-ended or differential input type for the analog input channel. Currently, the DT7837 supports only single-ended inputs.

If you want to read the value of the tachometer measurement in the input stream, you must also open and configure the tachometer subsystem; refer to page 39.

If you want to read the value of the counter/timer channel, you must open and configure the counter/timer subsystem; refer to page 41 for more information.

If you want to read the value of the measure counter, you must open and configure the measure counter subsystem; refer to page 48 for more information.

Note: The digital input port does not require configuration; therefore, you do not need to open or configure this subsystem.

You can return the current configuration of the analog input subsystem using the **ioctl** - **IOCTL_AIN_CFG_GET** command.

Configuring the Channel Mask for the Input Stream

Once you have configured the input channel for each subsystem that you want to use, set up the channel mask for the input stream using the **ioctl - IOCTL_CHAN_MASK_SET** command. You specify the channels that you want to enable for acquisition. Channels can include the analog input channels (bits 0 to 3 of the channel mask), tachometer (bit 8 of the channel mask), measure counter (bit 10 of the channel mask), and/or digital input port (bit 11 of the channel mask).

You can return the current configuration of the channel mask using the **ioctl - IOCTL_CHAN_MASK_GET** command.

Configuring the Sample Clock

Use the **ioctl - IOCTL_SAMPLE_CLK_SET** command to configure the sample clock for the input stream. For the DT7837, only the internal clock source is supported. You must specify the frequency at which to sample the specified channels in the input stream using the internal clock. Supported values range from 195.3125 Hz to 105.469 kHz.

Note: According to sampling theory (Nyquist Theorem), specify a frequency that is at least twice as fast as the input's highest frequency component. For example, to accurately sample a 20 kHz signal, specify a sampling frequency of at least 40 kHz to avoid aliasing.

Once the sample clock is started, the module requires 39 conversions before the first A/D sample is valid. The valid sample is aligned with the start trigger.

Note: The DT7837 module has two power modes: low power mode and high power mode. Low power mode is used when you specify a sampling frequency less than 52.734 kHz. High power mode is used when you specify a sampling frequency greater than or equal to 52.734 kHz.

If you change the power mode from low to high power or from high power to low power, and then configure the device, the module is self-calibrated. You may notice that it takes time after the device is configured to complete the calibration process.

The DT7837 driver sets the frequency of the internal clock as close as possible to the value that you specify. However, the value that you specify may not be the actual value that is set. To return the actual sample clock frequency that was set, use the **ioctl - IOCTL_SAMPLE_CLK_SET** command.

Configuring the Trigger that Starts Acquisition

Use the **ioctl - IOCTL_START_TRIG_CFG_SET** command to specify a start trigger that starts acquisition for the input stream. The DT7837 module supports the following sources for the start trigger:

- Software trigger (*trig_src_sw*) When you specify this trigger source, the trigger event occurs when you start the analog input operation using the ioctl IOCTL_START_SUBSYS command for the input stream, described on page 34.
- External digital (TTL) trigger (*trig_src_ext*)– When you specify this source, the trigger event occurs when the module detects a rising- or falling-edge transition on the specified general-purpose input pin on the Digital connector. Using software, you can specify which edge is active and which of the following pins of the Digital connector to use for the external trigger:
 - Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
 - Pin 2 corresponds to bit 1 of the digital input port (value 0x2).
 - Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
 - Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
 - Pin 5 corresponds to bit 4 of the digital input port (value 0x10).
 - Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
 - Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
 - Pin 8 corresponds to bit 7 of the digital input port (value 0x80).

Note: If you assigned a general-purpose input signal as an external trigger (or counter clock or gate input), you can read the value of the signal as you would any other digital input signal. Refer to page 52 for more information on reading digital input values.

- **Threshold trigger** (*trig_src_threshold*) When you specify this source, the trigger event occurs when the signal attached to a specified analog input channel rises above or falls below a user-specified threshold value. Using software, you specify the following parameters:
 - Edge Specify a rising-edge threshold trigger if you want to trigger when the signal rises above a threshold level, or a falling-edge threshold trigger if you want to trigger when the signal falls below a threshold level.
 - Threshold channel Specify any one of the analog input channels as the threshold input channel.
 - Threshold level Specify a value between ±10 V for a gain of 1 or ±1 V for a gain of 10 as the threshold level. Note that in software, this value must be entered as counts.

To convert volts to raw counts, use the following formulas:

<u>Gain of 1</u>: counts = ((volts/10.0f) * (1 << (24 - 1)))<u>Gain of 10</u>: counts = ((volts/1.0f) * (1 << (24 - 1))) To convert raw counts to volts (using two's complement data encoding), use these formulas:

```
Gain of 1:volts = ((counts * 10.0f)/(1 << (24 - 1)))Gain of 10:volts = ((counts * 1.0f)/(1 << (24 - 1)))
```

Note: The DT7837 driver sets the threshold level as close as possible to the value that you specify. However, the value that you specify may not be the actual value that is set. To return the actual threshold level that was set, use the **ioctl - IOCTL_START_TRIG_CNF_GET** command.

You can return the current configuration of the start trigger using the **ioctl** - **IOCTL_START_TRIG_CFG_GET** command.

Submitting I/O Requests

The DT7837 module works with the Linux AIO (Asynchronous Input/Output) model to submit I/O requests to the driver. The application can submit one or many requests from a thread. Submitting a request does not cause the thread to block, and instead the thread can proceed to do other computations and submit further requests to the device while the original request is in process. The application is expected to process completions and organize logical computations itself without depending on threads to organize the use of data.

For the DT7837, you must perform the following steps to set up and submit I/O requests to the driver for an input operation:

- **1.** Open an asynchronous I/O context using the Linux command **io_setup**.
- 2. Allocate an array of I/O control blocks (struct iocbs).

Each I/O control block is a structure with a number of parameters, one of which is the buffer used to store the output waveform or pattern.

For the DT7837, buffers must be on a 32-byte boundary and the length of each buffer must be a multiple of 32 bytes.

Each channel requires a value of four bytes. The buffer must contain a value for each channel that is enabled in the output stream.

3. Submit I/O requests to the asynchronous I/O context using io_submit.

Refer to your Linux documentation, such as Refer to the following website for more information about this model: https://code.google.com/p/kernel/wiki/AIOUserGuide, for more information on these commands and the AIO model.

Arming and Starting Continuous Operations

Once you have configured all the parameters for the subsystems used in the input stream, configured the input stream, and set up and submitted I/O requests, arm the input stream using the **ioctl - IOCTL_ARM_SUBSYS** command.

If an external trigger or threshold trigger was configured as the start trigger, the continuous input operation starts when the module detects the trigger condition.

If a software trigger was specified as the start trigger, you must explicitly start the continuous input operation using the **ioctl - IOCTL_START_SUBSYS** command. The operation starts immedately after the execution of this command.

When triggered, the DT7837 module simultaneously acquires data from all of the channels specified in the input stream.

Note: You can start a continuous operation on both the input and output streams simultaneously by specifying a non-zero value for the *pSimultaneous* variable of the **ioctl - IOCTL_START_SUBSYS** command.

Getting the Status of Acquisition

Using the **ioctl - IOCTL_ACQ_STATUS_GET** command, you can determine if the input operation has been armed or triggered, and whether the input FIFO is empty (contains no data) or is full.

Processing I/O Requests

As I/O requests are processed, events are generated. Your application must process the I/O requests as they are completed using the Linux command **io_getevents**. As each request is completed, it is up to the application to retrieve the data from the buffers and manage the I/O control blocks to ensure that buffers are available to be filled.

You can also set a timeout value for the **io_getevents** command so that the operation will stop if the request is not completed within a specified time.

Refer to your Linux documentation for more information on the io_getevents command.

Dealing with Input Buffers

The order of the data in the input buffer is as follows, assuming that all channels are enabled in the input stream:

- Analog input channels 0 through 3. Each analog input sample is a 32-bit, two's complement raw A/D value.
- Tachometer input. This is a 32-bit, unsigned value.

- Counter/timer. This is a 32-bit, unsigned value.
- Measure/counter. This is a 32-bit, unsigned value.
- Digital input port. This is 32-bit, unsigned value. The digital input data is in the least significant eight bits.

Refer to page 25 for more information on the channel mask for the input stream.

Enabling Buffer Error Reporting

By default, input stream buffer overrun are not reported to user programs. You can enable error reporting so that these errors are reported from the kernel device driver to the user program through signals.

To enable error reporting, perform the following steps:

- 1. Write your buffer error handler.
- 2. Open the input stream and then set the process ID or group ID that will receive the signal for input stream events. By default, SIGIO is the signal that is used to report buffer overrun errors. However, you can choose another signal, such as SIGUSR1, SIGUSR2, and so on, for reporting errors.

Note: If you want to enable buffer error reporting for both the input and output streams, it is recommended that you choose a different signal for each stream.

- **3.** Get all the status flags for the input stream and then set the asynchronous flag (FASYNC) for the input stream.
- **4.** (Optional) If you want to use a different signal (SIGUSR1, SIGUSR2, and so on) instead of the SIGIO signal, register the signal.
- 5. Register the signal handler for SIGIO or the signal that you register in step 4.

The following code show how to enable and handle buffer overrun errors from the input stream device file; refer to your Linux documentation for more information:

```
//include files
#define __USE_GNU (1)
#include <signal.h>
#include <fcntl.h>
//Step 1
//Write an error handler. Note that the signal handler is
//registered in Step 5 below
void buff_overrun_handler(int sig)
{
//your code here
}
```

```
//Step 2
//Open input stream file and set the process ID for the signal
//to use for reporting errors
int fd_in = open("/dev/dt7837-stream-in", O_RDWR);
if ((fcntl(fd_in, F_SETOWN, getpid())) < 0)</pre>
{
    //handle error
}
//Step 3
//Get the flags for the input stream and set the async flag (FASYNC)
int oflags = fcntl(fd_in, F_GETFL);
if ((fcntl(fd_in, F_SETFL, oflags | FASYNC)) < 0)
{
    //handle error
}
//Step 4
//This step is optional. By default buffer errors are reported
//using the signal SIGIO.
//To use a different signal, such as SIGUSR1, SIGUSR2, and so on,
//register the signal by specifying its number
if ((fcntl(fd_in, F_SETSIG, SIGUSR1)) < 0)</pre>
{
    //handle error
//Step 5
//Register a signal handler for the default signal, SIGIO, or the
//signal specified in Step 4
struct sigaction act;
memset (&act, 0, sizeof(act));
act.sa_handler = buff_overrun_handler; //signal handler
act.sa_flags = SA_NODEFER;
sigemptyset(&act.sa_mask);
if (sigaction(SIGUSR1, &act, NULL)) //Use SIGIO if you skip step 4
{
    //handle error
}
```

Stopping Continuous Operations

Once started, a continuous operation repeats continuously until you stop it with the **ioctl - IOCTL_STOP_SUBSYS** command.

This command stops an operation that was previously started with the **ioctl** - **IOCTL_START_SUBSYS** command (a software trigger was specified) or was started when the specified trigger condition was detected.

This command stops the DMA engine immediately and no further data is collected. Asynchronous I/O control blocks that were submitted using **io_submit** are still in the AIO queue and will not be completed. To cancel these control blocks, use **io_cancel**.

If you want to restart the operation for the input stream, you must rearm the input stream using **ioctl - IOCTL_ARM_SUBSYS**, and, if a software trigger is specified, restart the operation using **ioctl - IOCTL_START_SUBSYS**.

Cleaning up Resources

Once you are finished acquiring data, clean up the resources used by performing the following steps:

- **1.** Cancel any outstanding I/O requests using the Linux command **io_cancel**.
- 2. Destroy the asynchronous I/O context using the Linux command io_destroy.
- **3.** Close the analog input subsystem, the input stream, and any other open subsystems used in the input stream, using the **close** command, described on page 23.

Refer to your Linux documentation for more information on these commands.

Analog Output and Output Stream Operations

The DT7837 has one analog output channel. You can write a single value to the analog output channel using the analog output subsystem.

You can also update the analog output channel, and if desired, the digital output port continuously using the output stream.

This section describes these operations.

Performing a Single Value Operation

This section describes how to perform a single value analog output operation.

Opening the Subsystem

To perform a single value operation, open the file corresponding to the analog output subsystem (/dev/dt7837-aout) using the Linux command **open**, described on page 22.

Updating the Value of the Analog Output Channel

To update the analog output channel, perform a synchronous write operation using the Linux command **write**, specifying the file descriptor for the analog output subsystem, the variable that contains the value to write, and the size of the variable that contains the analog output value.

The write operation is blocking, in that it does not return until the value is written. The operation stops automatically once the value is written.

Closing the Subsystem

When finished, close the analog output subsystem using the **close** command, described on page 23.

Performing a Continuous Output Operation

This section describes how to perform a continuous output operation.

Opening the Output Stream

To perform a continuous analog output operation, open the file corresponding to the output stream (/dev/dt7837-stream-out) using the Linux command **open**, described on page 22.

Configuring the Channel Mask for the Output Stream

Set up the channel mask for the output stream using the **ioctl - IOCTL_CHAN_MASK_SET** command. You specify the channels that you want to enable for output. For the DT7837, you can include the analog output channel (bit 16 of the channel mask) and/or the individual lines of the digital output port (bits 24 to 31 of the channel mask) in the output stream.

You can return the current configuration of the channel mask using the **ioctl - IOCTL_CHAN_MASK_GET** command.

Configuring the Sample Clock

Use the **ioctl - IOCTL_SAMPLE_CLK_SET** command to configure the clock source for the output stream. For the DT7837, only the internal clock is supported. You must specify the frequency at which to update the specified channels in the output stream using the internal clock. For the DT7837, supported values range from 30 kHz to 216 kHz.

The DT7837 module requires 36 sample clock periods once the output subsystem has been triggered before the first D/A conversion is completed. The clock is always running.

Note: The sample frequencies for the A/D and D/A subsystems on the DT7837 are independently programmable and are derived from the same 48 MHz reference clock. Therefore, it is possible to establish a fixed relationship between the A/D and D/A sample frequencies, including setting them to the same frequency (30 kHz to 105.469 kHz). However, the subsystems are not designed to be synchronous with each other.

The DT7837 driver sets the frequency of the internal clock as close as possible to the value that you specify. However, the value that you specify may not be the actual value that is set. To return the actual sample clock frequency that was set, use the **ioctl - IOCTL_SAMPLE_CLK_SET** command.

Configuring the Trigger that Starts the Output Operation

Use the **ioctl - IOCTL_START_TRIG_CFG_SET** command to specify a start trigger that starts the continuous operation for the output stream. The DT7837 module supports the following sources for the start trigger:

- Software trigger (*trig_src_sw*) When you specify this trigger source, the trigger event occurs when you start the analog output operation using the **ioctl IOCTL_START_SUBSYS** command for the output stream, described on page 34.
- External digital (TTL) trigger (*trig_src_ext*)– When you specify this source, the trigger event occurs when the module detects a rising- or falling-edge transition on the specified general-purpose input pin on the Digital connector. Using software, you can specify which edge is active and which of the following pins of the Digital connector to use for the external trigger:
 - Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
 - Pin 2 corresponds to bit 1 of the digital input port (value 0x2).

- Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
- Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
- Pin 5 corresponds to bit 4 of the digital input port (value 0x10).
- Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
- Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
- Pin 8 corresponds to bit 7 of the digital input port (value 0x80).

You can return the current configuration of the start trigger using the **ioctl - IOCTL_START_TRIG_CFG_GET** command.

Submitting I/O Requests

The DT7837 module works with the Linux AIO (Asynchronous Input/Output) model to submit I/O requests to the driver. The application can submit one or many requests from a thread. Submitting a request does not cause the thread to block, and instead the thread can proceed to do other computations and submit further requests to the device while the original request is in process. The application is expected to process completions and organize logical computations itself without depending on threads to organize the use of data.

For the DT7837, you must perform the following steps to set up and submit I/O requests to the driver for an output operation:

- 1. Open an asynchronous I/O context using the Linux command io_setup.
- 2. Allocate an array of I/O control blocks (struct iocbs).

Each I/O control block is a structure with a number of parameters, one of which is the buffer used to store the data that is acquired.

For the DT7837, buffers must be on a 32-byte boundary and the length of each buffer must be a multiple of 32 bytes. Each sample requires four bytes.

- **3.** Fill each buffer with the pattern that you want to output to the enabled channels in the output stream.
- 4. Submit I/O requests to the asynchronous I/O context using io_submit.

Refer to your Linux documentation, such as Refer to the following website for more information about this model: https://code.google.com/p/kernel/wiki/AIOUserGuide, for more information on these commands and the AIO model.

Arming and Starting Continuous Operations

Once you have configured the output stream and set up and submitted I/O requests, arm the output stream using the **ioctl - IOCTL_ARM_SUBSYS** command.

If an external trigger was configured as the start trigger, the continuous output operation starts when the module detects the trigger condition.

If a software trigger was specified as the start trigger, you must explicitly start the continuous input operation using the **ioctl - IOCTL_START_SUBSYS** command. The operation starts immedately after the execution of this command.

When triggered, the DT7837 module simultaneously updates all of the channels specified in the output stream.

Note: You can start a continuous operation on both the input and output streams simultaneously by specifying a non-zero value for the *pSimultaneous* variable of the **ioctl - IOCTL_START_SUBSYS** command.

Getting the Status of the Output Operation

Using the **ioctl - IOCTL_ACQ_STATUS_GET** command, you can determine if the output operation has been armed or triggered, and whether the output FIFO is empty (contains no data) or is full.

Processing I/O Requests

As I/O requests are processed, events are generated. Your application must process the I/O requests as they are completed using the Linux command **io_getevents**. As each request is completed, it is up to the application to retrieve the data from the buffers and manage the I/O control blocks to ensure that buffers are filled and available.

You can also set a timeout value for the **io_getevents** command so that the operation will stop if the request is not completed within a specified time.

Refer to your Linux documentation for more information on the **io_getevents** command.

Dealing with Output Buffers

The order of the data in the output buffer is as follows, assuming that all channels are enabled in the output stream:

- Analog output channel 0. The analog output sample is a 32-bit, two's complement value.
- Digital output port. This is 32-bit unsigned value. The digital output data is in the least significant eight bits.

Refer to page 33 for more information on the channel mask for the output stream.

Enabling Buffer Error Reporting

By default, output stream buffer underrun errors are not reported to user programs. You can enable error reporting so that these errors are reported from the kernel device driver to the user program through signals.

To enable error reporting for the output stream, perform the following steps:

1. Write your buffer error handler.

2. Open the output stream and then set the process ID or group ID that will receive the signal for output stream events. By default, SIGIO is the signal that is used to report buffer overrun errors. However, you can choose another signal, such as SIGUSR1, SIGUSR2, and so on, for reporting errors.

Note: If you want to enable buffer error reporting for both the input and output streams, it is recommended that you choose a different signal for each stream.

- **3.** Get all the status flags for the output stream and then set the asynchronous flag (FASYNC) for the output stream.
- **4.** (Optional) If you want to use a different signal (SIGUSR1, SIGUSR2, and so on) instead of the SIGIO signal, register the signal.
- 5. Register the signal handler for SIGIO or the signal that you register in step 4.

The following code show how to enable and handle buffer underrun errors from the output stream device file.

```
//include files
#define ___USE_GNU
                     (1)
#include <signal.h>
#include <fcntl.h>
//Step 1
//Write an error handler. Note that the signal handler is
//registered in Step 5 below
void buff_underrun_handler(int sig)
{
//your code here
}
//Step 2
//Open output stream file and set the process ID for the signal
//to use for reporting errors
int fd_out = open("/dev/dt7837-stream-out", O_RDWR);
if ((fcntl(fd_out, F_SETOWN, getpid())) < 0)</pre>
{
    //handle error
}
//Step 3
//Get the flags for the output stream and set the async flag (FASYNC)
int oflags = fcntl(fd_out, F_GETFL);
if ((fcntl(fd_out, F_SETFL, oflags | FASYNC)) < 0)
{
    //handle error
}
```
Using the File I/O Commands

```
//Step 4
//This step is optional. By default buffer errors are reported
//using the signal SIGIO.
//To use a different signal, such as SIGUSR1, SIGUSR2, and so on,
//register the signal by specifying its number
if ((fcntl(fd_out, F_SETSIG, SIGUSR1)) < 0)</pre>
{
    //handle error
}
//Step 5
//Register a signal handler for the default signal, SIGIO, or the
//signal specified in Step 4
struct sigaction act;
memset (&act, 0, sizeof(act));
act.sa_handler = buff_underrun_handler; //signal handler
act.sa_flags = SA_NODEFER;
sigemptyset(&act.sa_mask);
if (sigaction(SIGUSR1, &act, NULL)) //Use SIGIO if you skip step 4
{
    //handle error
}
```

Stopping Continuous Operations

Once started, a continuous operation repeats continuously until you stop it with the **ioctl** - **IOCTL_STOP_SUBSYS** command.

This command stops an operation that was previously started with the **ioctl** - **IOCTL_START_SUBSYS** command (a software trigger was specified) or was started when the specified trigger condition was detected.

This command stops the DMA engine immediately and no further data is output. Asynchronous I/O control blocks that were submitted using **io_submit** are still in the AIO queue and will not be completed. To cancel these control blocks, use **io_cancel**.

If you want to restart the operation for the output stream, you must rearm the output stream using **ioctl - IOCTL_ARM_SUBSYS**, and, if a software trigger is specified, restart the operation using **ioctl - IOCTL_START_SUBSYS**.

Cleaning up Resources

Once you are finished outputting data, clean up the resources used by performing the following steps:

- **1.** Cancel any outstanding I/O requests using the Linux command **io_cancel**.
- 2. Destroy the asynchronous I/O context using the Linux command io_destroy.
- 3. Close the output stream using the close command, described on page 23.

Refer to your Linux documentation for more information on these commands.

Tachometer Operations

The DT7837 module supports one tachometer input signal. You can measure the frequency or period of the tachometer input signal to calculate the rotation speed for high-level (\pm 30 V) tachometer input signals. An internal 12 MHz counter is used for the measurement, yielding a resolution of 83 ns (1/12 MHz). The value of the tachometer can be returned in the input stream.

To read the tachometer measurement in the input stream, set bit 8 of the channel mask of the input stream. Refer to page 32 for more information on analog input and input stream operations.

This section describes how to configure the parameters of the tachometer subsystem.

Opening the Subsystem

Open the file corresponding to the tachometer subsystem (/dev/dt7837-tach using the Linux command **open**, described on page 22.

Configuring the Tachometer Subsystem

To configure the tachometer subsystem, use the **ioctl - IOCTL_TACH_CFG_SET** command.

You specify the file descriptor for the tachometer subsystem and the following parameters:

- The edge (rising or falling) of the tachometer to use for the measurement. The number of counts betweeen two consecutive edges of the tachometer input signal is used as the tachometer measurement.
- The Stale flag (*stale_flag*) that indicates whether or not the data is new. If *stale_flag* is set as Used (1), the most significant bit (MSB) of the value is set to 0 to indicate new data; reading the value before the measurement is complete returns an MSB of 1. If *stale_flag* is set to Not Used (0), the MSB is always set to 0.

When the input operation is started, the internal 12 MHz counter starts incrementing when it detects the first starting edge of the tachometer input and stops incrementing when it detects the next starting edge; at that point, the counter stores the count. The stored count is maintained until it is read as part of the input data stream or until a new count is stored. The next tachometer measurement operation is started automatically.

If the sample rate of the input subsystem is faster than the tachometer input frequency, then the stored count retains the current value when the count is read by the input subsystem. The operation of *stale_flag* in this case is described as follows:

- If another input subsystem sample occurs before another measure completes and *stale_flag* is used, then the Stale flag is set and the stale measure count is written into the input data stream.
- If another input subsystem sample occurs before another measure completes and *stale_flag* is not used, then the Stale flag is not set and the stale measure count is written into the input data stream.

If the input sample rate is slower than the tachometer input frequency, then as each period measurement completes, a new count value is stored. When the input subsystem sample occurs, the most recently stored measure count is written into the input data stream.

A data pipeline is used in the hardware to compensate for the A/D group delay and synchronizes the value of the tachometer input with the analog input measurements so that all measurements are correlated in time.

When you read the value of the tachometer input as part of the input stream, you might see results similar to the following:

Time	A/D Value	Tachometer Input Value	Status of Operation
10	5002	0	Operation started, but is not complete
20	5004	0	Operation not complete
30	5003	0	Operation not complete
40	5002	12373	Operation complete
50	5000	12373	Next operation started, but is not complete
60	5002	12373	Operation not complete
70	5004	12373	Operation not complete
80	5003	14503	Operation complete
90	5002	14503	Next operation started, but is not complete

Table 3: An Example of Reading the Tachometer Input as Part of the Input Stream

Using the count that is returned from the tachometer input, you can determine the following:

- Frequency of a signal pulse (the number of periods per second). You can calculate the frequency as follows:
 - Frequency = 12 MHz/(Number of counts 1) where 12 MHz is the internal counter/timer clock frequency

For example, if the count is 21, the measured frequency is 600 kHz (12 MHz/20).

- Period of a signal pulse. You can calculate the period as follows:
 - Period = 1/Frequency
 - Period = (Number of counts 1)/12 MHz where 12 MHz is the internal counter/timer clock frequency

You can return the current configuration of the tachometer subsystem using the **ioctl - IOCTL_TACH_CFG_GET** command.

Counter/Timer Operations

The DT7837 module supports one counter/timer channel. You can read the value of the counter directly using a synchronous read of the counter/timer subsystem.

This section describes counter/timer operations.

Opening the Subsystem

To perform a synchronous read of the counter/timer, open the file corresponding to the counter/timer subsystem (/dev/dt7837-ctr-tmr) using the Linux command **open**, described on page 22.

Configuring the Counter/Timer Subsystem

To configure the counter/timer subsystem, use the **ioctl - IOCTL_CT_CFG_SET** command.

You specify the file descriptor for the counter/timer subsystem and then specify the parameters specific to the counter/timer subsystem. This section describes each of the configurable parameters of the counter/timer subsystem.

You can return the current configuration of the counter/timer subsystem using the **ioctl - IOCTL_CT_CFG_GET** command.

Mode

The DT7837 module supports the following counter/timer modes:

- Event counting described below.
- Rate generation described on page 42
- Non-retriggerable one-shot mode described on page 43
- Idle described on page 44.

Event Counting Mode

Use event counting mode (*ct_mode_counter*) if you want to count the number of rising edges that occur on the counter's clock input when the counter's gate signal is active (low-level or high-level).

You can count a maximum of 4,294,967,296 events before the counter rolls over to 0 and starts counting again.

Using software, you must specify the following parameters for the event counting operation:

- Active gate type (external low level or external high level). Refer to page 44 for more information about the supported gate types.
- The general-purpose input pin (1 to 8) of the Digital connector to use for the external gate signal. Ensure that you physically connect the gate signal to this input pin. Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.
- The C/T clock source (internal or external). Note that in event counting mode, the external C/T clock is more useful than an internal C/T clock; refer to page 45 for more information about the C/T clock sources.
- The general-purpose input pin (1 to 8) of the Digital connector to use for the external C/T clock input. Ensure that you physically connect the clock input signal to this input pin. Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.

Rate Generation Mode

Use rate generation mode (*ct_mode_divider*) to generate a continuous pulse output signal from the counter's output signal. You can use this pulse output signal as an external clock to pace other operations, such as an analog input or other counter/timer operations.

The pulse output operation is enabled whenever the counter's gate signal is active. While the pulse output operation is enabled, the counter outputs a pulse of the specified type and frequency continuously. As soon as the operation is disabled, rate generation stops.

You can output pulses using a maximum frequency of 24 MHz (if using the internal C/T clock) or 5 MHz (if using the external C/T clock).

Note: The integrity of the signal degrades at frequencies greater than 10 MHz.

Using software, you must specify the following parameters for the rate generation operation:

- Active gate type (external low level or external high level). Refer to page 44 for more information about the supported gate types.
- The general-purpose input pin (1 to 8) of the Digital connector to use for the external gate signal. Ensure that you physically connect the gate signal to this input pin. Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.
- The C/T clock source (internal or external). Refer to page 45 for more information about the C/T clock sources.
- If you are using an external C/T clock source, the general-purpose input pin (1 to 8) of the Digital connector to use for the external C/T clock input. Ensure that you physically connect the clock input signal to this input pin. Refer to *DT7837 User's Manual* for the pin descriptions of the Digital connector.
- The period of the output pulse. Refer to page 46 for more information about the period of the output pulse.
- The pulse width (duty cycle) of the active pulse. Refer to page 46 for more information about the pulse width of the output pulse.

- The polarity of the output signal (active high or active low). Refer to page 46 for more information on the polarity of the output pulse.
- The general-purpose output pin (11 to 18) of the Digital connector to use for the external C/T clock output signal. Ensure that you physically connect the C/T output signal to this output pin. Refer to *DT7837 User's Manual* for the pin descriptions of the Digital connector.

Non-Retriggerable One-Shot Mode

Use non-retriggerable one-shot mode (*ct_mode_1shot*) to generate a single output pulse from the counter whenever the specified edge is detected on the counter's gate signal (after the pulse period from the previous output pulse expires). Any gate signals that occur while the pulse is being output are not detected by the module, as shown in Figure 3. The module continues to output a pulse when the specified gate edge is detected until you stop the operation. You can use this mode to clean up a poor clock input signal by changing its pulse width, and then outputting it.



*You can determine period, pulse width, and polarity of the output pulse using software.



Using software, you must specify the following additional parameters to set up the non-retriggerable one-shot operation:

- Active gate type that enables the output pulse. Refer to page 44 for more information about the supported gate types.
- The general-purpose input pin (1 to 8) of the Digital connector to use for the external gate signal. Ensure that you physically connect the gate signal to this input pin. Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.
- The C/T clock source (internal or external) used to generate the pulse. Note that in non-retriggerable one-shot mode, the internal C/T clock is more useful than an external C/T clock; refer to page 45 for more information about the C/T clock sources.

- If using the external C/T clock source, the general-purpose input pin (1 to 8) of the Digital connector to use for the external C/T clock input. Ensure that you physically connect the clock input signal to this input pin. Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.
- The period of the output pulse. Refer to page 46 for more information about the period of the output pulse.
- The pulse width (duty cycle) of the output pulse. Refer to page 46 for more information about the pulse width of the output pulse.
- The general-purpose output pin (11 to 18) of the Digital connector to use for the external C/T clock output signal. Ensure that you physically connect the C/T output signal to this output pin. Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.
- Retriggerable setting of 0 for non-retriggerable one-shot mode.
- The polarity of the output signal (active high or active low). Refer to page 46 for more information on the polarity of the output pulse.

Idle Mode

When you use idle mode (*ct_mode_idle*), the counter no longer drives the clock output signal that is assigned to one of the general-purpose output signals (pins 11 to 18) of the Digital connector.

Note: The value of the counter output signal can also be overwritten by writing to the digital output subsystem. Refer to page 53 for more information.

If you assigned a general-purpose input signal as a counter clock or gate input (or external trigger), you can read the value of the signal as you would any other digital input signal. Refer to page 52 for more information on reading digital input values.

Gate

The counter's gate signal determines when a counter/timer operation is enabled.

DT7837 modules provide the following choices for the gate parameter:

- None (*ct_gate_none*) A software start command enables the counter/timer operation immediately after execution. (No general-purpose input signal is required if a gate type of *ct_gate_none* is selected.)
- Low external gate input (*ct_gate_ext_lo*) Specifies a logic low or falling edge gate type. For event counting and rate generation mode, the operation is enabled when the counter's gate signal is low and is disabled when the counter's gate signal is high. For one-shot or repetitive one-shot mode, the operation is enabled when the counter's gate signal goes from a high to a low transition and is disabled when the counter's gate signal goes from a low transition.

Using software, you specify one of the general-purpose input pins of the Digital connector on the DT7837 module as the external C/T gate input (ext_gate_din). Ensure that you physically connect the external gate signal to the selected pin. (Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.)

• **High external gate input** (*ct_gate_ext_hi*) – Specifies a logic high or rising edge gate type. For event counting and rate generation mode, the operation is enabled when the counter's gate signal is high and is disabled when the counter's gate signal is low. For one-shot mode and repetitive one-shot mode, the operation is enabled when the counter's gate signal goes from a low to a high transition and is disabled when the counter's gate signal goes from a high to a low transition.

You specify one of the general-purpose input pins of the Digital connector on the DT7837 module as the external C/T gate input (ext_gate_din). Ensure that you physically connect the external gate signal to the selected pin. (Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.)

C/T Clock Input Sources

The following input clock sources are available for the general-purpose counter/timer:

- Internal C/T clock The internal C/T clock on the DT7837 uses a 48 MHz time base. This clock source is typically used for one-shot, repetitive one-shot, and rate generation operations.
- External C/T clock An external C/T clock is useful when you want to pace counter/timer operations at rates not available with the internal C/T clock or if you want to pace at uneven intervals. The frequency of the external C/T clock can range from 0.0112 Hz to 10 MHz.

This clock source is typically used for event counting operations or rate generation operations.

You specify one of the general-purpose input pins (1 to 8) of the Digital connector on the DT7837 module as the external C/T clock (ext_clk_din). Then, physically connect the external clock signal to the selected general-purpose input pin. (Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.) Counter/timer operations start on the rising edge of the clock input signal.

Note: If you specify a counter/timer in the input stream, the A/D sample clock determines how often you want to read the counter value. Refer to page 33 for more information about the A/D sample clock.

Pulse Output Period, Pulse Width, and Polarity

If you want to perform a C/T output operation, use software to define one of the general-purpose output pins (11 to 18) of the Digital connector on the DT7837 module as the external C/T output signal. Then, connect the external C/T output signal to the selected general-purpose output pin. (Refer to the *DT7837 User's Manual* for the pin descriptions of the Digital connector.)

For the DT7837 module, you can program the polarity of the output pulse (active high or active low). For an active high pulse, the high portion of the total pulse output period is the active portion of the counter/timer pulse output signal. For an active low pulse, the low portion of the total pulse output period is the active portion of the counter/timer pulse output signal.

You can specify the number of input clock cycles that are used to create one period of the counter clock output signal. You can also specify the number of input clock cycles used to create the active pulse width (or duty cycle) of the C/T output signal.

For example, if you are using an external C/T clock running at 10000 Hz as the input clock source of the counter/timer, and you want to generate a output signal of 1000 Hz with a 20% duty cycle, specify a period of 10 (10000 Hz divided by 10 is 1000 Hz) and a pulse width of 2 (the period of 10 multiplied by 20%). This is illustrated in Figure 4.



Figure 4: Example of a Pulse Output

Note: If you are using an internal C/T clock input source, you can output pulses using a maximum frequency of 24 MHz. Note, however, that the integrity of the signal degrades at frequencies greater than 10 MHz.

If you are using an external C/T clock input source, you can output pulses using a maximum frequency of 5 MHz.

Starting the Counter/Timer Operation

Once you have configured all the parameters for the counter/timer subsystem, start the counter/timer using the **ioctl - IOCTL_START_SUBSYS** command. If a software gate is specified, the operation starts immedately after the execution of this command. For other gate types, the operation starts when the specified gate type is enabled.

Reading the Counter/Timer

You can read the value of the counter using the Linux command **read**. You specify the file descriptor for the counter/timer subsystem, the variable that will store the resulting value, and the size of the variable that will store the counter value.

Stopping the Counter/Timer Operation

To stop the counter/timer operation, use the **ioctl - IOCTL_STOP_SUBSYS** command. When this command is executed, the counter/timer operation stops immediately.

Closing the Subsystem

When you are finished reading the value of the counter/timer, close the counter/timer subsystem using the **close** command, described on page 23.

Measure Counter Operations

The DT7837 module supports one measure counter whose value can be returned in the input stream. Using this counter, you can measure the frequency, period, or pulse width of a single signal or the time period between two signals. The measure counter is useful for correlating analog input data with digital positional data, measuring the frequency of a signal, or as a tachometer. An internal 48 MHz counter is used for the measurement, yielding a resolution of 20.83 ns (1/48 MHz).

The read the value of the measure counter in the input stream, set bit 10 of the channel mask of the input stream. Refer to page 25 for more information on analog input and input stream operations.

This section describes how to configure the parameters of the measure counter.

Opening the Subsystem

Open the file corresponding to the measure counter subsystem (/dev/dt7837-measure) using the Linux command **open**, described on page 22.

Configuring the Measure Counter Subsystem

To configure the measure counter subsystem, use the **ioctl - IOCTL_MCTR_CFG_SET** command.

You specify the file descriptor for the measure counter subsystem and then specify the following parameters:

• The signals that start and stop the measurement. Refer to Table 4 for the supported start and stop signals.

Signal	Connection Required
A/D conversion complete	No connection required.
Tachometer input (falling edge or rising edge)	Connect to the Tachometer input (pin 23).
Digital input 0 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 0 (pin 1) of the Digital connector. By default, this is digital input 0.
Digital input 1 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 1 (pin 2) of the Digital connector. By default, this is digital input 1.
Digital input 2 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 2 (pin 3) of the Digital connector. By default, this is digital input 2.

Table 4:	Possible	Start and	Stop	Signals
----------	----------	-----------	------	---------

Signal	Connection Required
Digital input 3 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 3 (pin 4) of the Digital connector. By default, this is digital input 3.
Digital input 4 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 4 (pin 5) of the Digital connector. By default, this is digital input 4.
Digital input 5 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 5 (pin 6) of the Digital connector. By default, this is digital input 5.
Digital input 6 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 6 (pin 7) of the Digital connector. By default, this is digital input 6.
Digital input 7 (falling edge or rising edge)	Connect a digital input, external A/D trigger, C/T clock input, or C/T gate input to general-purpose input 7 (pin 8) of the Digital connector. By default, this is digital input 7.

Table 4: Possible Start and Stop Signals

• A Stale flag (*stale_flag*) indicating whether or not the data is new. This flag is used only when the start edge and the stop edge is set to use the same pin and edge (such as pin 0 - DIN 0 rising as the start edge and pin 0 -DIN rising as the stop edge).

If *stale_flag* is set as Used (1), the most significant bit (MSB) of the value is set to 0 to indicate new data; reading the value before the measurement is complete returns an MSB of 1. If *stale_flag* is set to Not Used (0), the MSB is always set to 0.

When the selected start edge is the same as the selected stop edge, the internal 48 MHz counter starts incrementing when it detects the first start edge of the selected input signal and stops incrementing when it detects the selected stop edge (which is the same as the start edge, in this case); at that point, the counter stores and resets the count. The stored count is maintained until it is read as part of the input data stream or until a new count is stored. Since the stop edge is the same as the start edge in this case, the stop edge for the current measurement is the start edge for the next measurement; therefore, no waveform periods are missed. The value of the measure count depends on the input subsystem sample frequency, described as follows:

- If the input subsystem sample frequency is faster than the selected input frequency, then the stored measure count retains the current value when it is read by the input subsystem. The operation of the Stale flag in this case is described as follows:
 - If another input subsystem sample occurs before another measure completes and the Stale flag is used, then the Stale flag is set and the stale measure count is written into the input data stream.
 - If another input subsystem sample occurs before another measure completes and the Stale flag is not used, then the Stale flag is not set and the stale measure count is written into the input data stream.

• If the input subsystem sample frequency is slower than the selected input frequency, then the new measure count value is stored as each period measurement completes. When an input subsystem sample occurs, then the most recently stored measure count is written into the input data stream.

When the selected start edge is not the same as the selected stop edge, the internal 48 MHz counter starts incrementing when it detects the selected start edge and stops incrementing when it detects the next selected stop edge; at that point, the counter stores and resets the count. The stored count is maintained until it is read as part of the input data stream or until a new count is stored. The value of the measure count depends on the input subsystem sample frequency, described as follows:

- If the input subsystem sample rate is faster than the selected measurement period, then the stored count retains the current value when the count is read by the input subsystem. The operation of *stale_flag* in this case is described as follows:
 - If another input subsystem sample occurs before another measure completes and *stale_flag* is used, then *stale_flag* is set and the stale measure count is written into the input data stream.
 - If another input subsystem sample occurs before another measure completes and *stale_flag* is not used, then *stale_flag* is not set and the stale measure count is written into the input data stream.
- If the input subsystem sample rate is slower than the selected measurement period, then a new count value is stored as each period measurement completes. When an input subsystem sample occurs, the most recently stored measure count is written into the input data stream.

A data pipeline is used in the hardware to compensate for the A/D group delay and synchronizes the value of the measure counter with the analog input measurements, so that all measurements are correlated in time.

When you read the value of the measure counter as part of the input data stream, you might see results similar to the following:

Time	A/D Value	Measure Counter Values	Status of Operation
10	5002	0	Operation started, but is not complete
20	5004	0	Operation not complete
30	5003	0	Operation not complete
40	5002	12373	Operation complete
50	5000	12373	Next operation started, but is not complete
60	5002	12373	Operation not complete

Table 5: An Example of Reading a Measure Counter as Part of the Input Stream

Time	A/D Value	Measure Counter Values	Status of Operation
70	5004	12373	Operation not complete
80	5003	14503	Operation complete
90	5002	14503	Next operation started, but is not complete

Table 5: An Example of Reading a Measure Counter as Part of the Input Stream (cont.)

Using the count that is returned from the measure counter, you can determine the following:

- Frequency between the start and stop signals/edges. You can calculate the frequency as follows:
 - Frequency = 48 MHz/(Number of counts 1) where 48 MHz is the internal measure counter frequency

For example, if the count is 201, the measured frequency is 240 kHz (48 MHz/200).

- Period between the start and stop signals/edges. You can calculate the period as follows:
 - Period = 1/Frequency
 - Period = (Number of counts 1)/48 MHz where 48 MHz is the internal measure counter frequency
- Pulse width of the start and stop signal/edges (rising to falling edge or falling edge to rising edge). You can calculate the period as follows:
 - Pulse width = 1/Frequency
 - Pulse width = (Number of counts 1)/48 MHz where 48 MHz is the internal measure counter frequency

You can return the current configuration of the measure counter subsystem using the **ioctl - IOCTL_MCTR_CFG_GET** command.

Digital Input Operations

The DT7837 module supports synchronous reads of the digital input port. You can also read the value of the digital port in the input stream by setting bit 11 of the channel mask of the input stream; refer to page 25 for more information on input stream operations.

The digital input subsystem does not need to be configured. By default, general-purpose input pins 1 to 8 of the Digital connector on the DT7837 module correspond to digital input signals 0 to 7.

Note: If you assigned a general-purpose input signal as a counter clock or gate input or as an external trigger, you can read the value of the signal as you would any other digital input signal, if desired.

A digital line is high if its value is 1; a digital line is low if its value is 0. On power up or reset, a low value (0) is output from each of the digital output lines and a high value (1) is read from each of the digital input lines if the lines are not connected.

This section describes synchronous read operations of the digital input port.

Opening the Subsystem

Open the file corresponding to the digital input subsystem (/dev/dt7837-din) using the Linux command **open**, described on page 22.

Reading the Value

To perform a synchronous read of the digital input port, use the Linux command **read**, specifying the file descriptor for the digital input subsystem, the variable that will store the resulting value, and the size of the variable that will store the digital input value.

The read operation is blocking, in that it does not return until the value is read. The operation stops automatically once the value is returned.

Closing the Subsystem

When finished, close the digital input subsystem using the **close** command, described on page 23.

Digital Output Operations

The DT7837 module supports synchronous writes to the digital output port. You can also update the individual lines of the digital output port through the output stream by setting bits 24 to 31 of the channel mask for the output stream; refer to page 33 for more information on output stream operations.

The digital output subsystem does not need to be configured. By default, general-purpose input pins 11 to 18 of the Digital connector on the DT7837 module correspond to digital output signals 0 to 7.

Note: If the clock output signal of the counter/timer was assigned to one of the general-purpose output signals, you can overwrite the value of the signal by writing to the digital output subsystem. Therefore, ensure that you know the configuration of each output pin of the Digital connector before writing to it or you could corrupt the signal on that pin.

A digital line is high if its value is 1; a digital line is low if its value is 0. On power up or reset, a low value (0) is output from each of the digital output lines and a high value (1) is read from each of the digital input lines if the lines are not connected.

This section describes synchronous write operations of the digital output port.

Opening the Subsystem

Open the file corresponding to the digital output subsystem (/dev/dt7837-dout) using the Linux command **open**, described on page 22.

Updating the Value of the Port

To perform a synchronous write to the digital output port, use the Linux command **write**, specifying the file descriptor for the digital output subsystem, the variable that contains the value to write, and the size of the variable that contains the digital output value.

The write operation is blocking, in that it does not return until the value is written. The operation stops automatically once the value is written.

Closing the Subsystem

When finished, close the digital output subsystem using the **close** command, described on page 23.

Calibration

This section describes the calibration process for the analog input and analog output circuitry on the DT7837.

Note: DT7837 modules are calibrated at the factory and should not require calibration for initial use. It is recommended that you check and, if necessary, readjust the calibration of the analog circuitry every six months using the DT7837 calibration program. Refer to the *DT7837 User's Manual* for more information on this program. This section describes the commands that the DT7837 calibration program uses to calibrate the module. You can write your own calibration program, if desired, using these commands.

Analog Input Calibration

Each analog input channel can be calibrated for offset (zero) and gain to compensate for variances in the Analog-to-Digital Converter (ADC) and associated circuitry. To do this, each analog input channel has an associated offset and gain potentiometer for each gain setting and for sampling frequencies less than 52.734 kHz and greater than or equal to 52.734 kHz. As shown in Table 6, user calibration registers correspond to different gain and sampling frequency combinations.

Analog Input Calibration Registers	Gain	Sampling Frequency
0	1	> 52.734 kHz
1	10	> 52.734 kHz
2	1	< = 52.734 kHz
3	10	< = 52.734 kHz

Table 6: Analog Input Calibration Registers

Each potentiometer has three values: a "factory" calibration value that is programmed at the factory and represents a known accurate configuration, a "user" calibration value that you can modify, and a "wiper" value, which is the current value. In most cases, the wiper and user calibration values are the same. They differ only during the actual calibration process when the wiper value is modified until the desired value is reached.

Note: The DT7837 module has two power modes: low power mode and high power mode. Low power mode is used when you specify a sampling frequency less than 52.734 kHz. High power mode is used when you specify a sampling frequency greater than or equal to 52.734 kHz.

If you change the power mode from low to high power or from high power to low power, and then configure the device, the module is self-calibrated. You may notice that it takes time after the device is configured to complete the calibration process.

Also, if you change the sampling rate, you must wait a few milliseconds for the master clock to settle before performing calibrating the module.

This section describes how to calibrate the analog input circuitry.

Opening the Subsystem

Open the file corresponding to the analog input subsystem (/dev/dt7837-ain) using the Linux command **open**, described on page 22.

Calibrating the Offset

To calibrate the offset of an analog input channel, connect a precision voltage source of 0 V to the specified analog input channel. Then, use the **ioctl - IOCTL_OFFSET_POT_SET** command to specify the analog input channel that you want to calibrate, the value to write to the potentiometer, whether the calibration type is a user (0) or factory (1) calibration, and the calibration register to update (0 to 3) based on the gain and the sampling frequency used.

Adjust the potentiometer until you measure a value as close as possible to 0 V. You can continue to adjust the offset values by calling this command with different values. The offset calibration register is automatically written to the wiper of the associated potentiometer.

When you are satisfied that the offset calibration value is correct, call the **ioctl** - **IOCTL_OFFSET_POT_WIPER_SET** command. You specify the analog input channel (potentiometer) that you want to calibrate and the value to write to the potentiometer. The value is written to a calibration register in non-volatile EEPROM.

These steps should be repeated until all the offset calibration registers (for all gains and sampling frequencies) are calibrated for each analog input channel.

Note that you can read the value of the offset potentiometers using the **ioctl** - **IOCTL_OFFSET_POT_GET** and **ioctl** - **IOCTL_OFFSET_POT_WIPER_GET** commands.

Calibrating the Gain

To calibrate the gain of an analog input channel, connect a precision voltage source of +9.375 V to the specified analog output channel.

Then, use the **ioctl - IOCTL_GAIN_POT_SET** command to specify the analog input channel that you want to calibrate, the value to write to the potentiometer, whether the calibration type is a user or factory calibration, and the calibration register to update (0 to 3) based on the gain and sampling frequency used.

Note: This command will block all other operations for at least 5 ms.

You can continue to adjust the gain calibration values by calling this command with different values. The gain calibration register is automatically written to the wiper of the associated potentiometer.

When you are satisfied that the gain calibration value is correct, call the **ioctl -IOCTL_GAIN_POT_WIPER_SET** command. You specify the analog input channel (potentiometer) that you want to calibrate and the value to write to the potentiometer. The value is written to a calibration register in non-volatile EEPROM.

These steps should be repeated until all the gain calibration registers (for all gains and sampling frequencies) are calibrated for each analog input channel.

Note that you can read the value of the gain potentiometers using the **ioctl** - **IOCTL_GAIN_POT_GET** and **ioctl** - **IOCTL_GAIN_POT_WIPER_GET** commands.

Closing the Subsystem

When finished, close the analog input subsystem using the **close** command, described on page 23.

Analog Output Calibration

Each analog output channel can be calibrated for offset (zero) and gain to compensate for variances in the Digital-to-Analog Converter (DAC) and associated circuitry. To do this, each analog output channel has an associated offset and gain potentiometer for three output frequency ranges, as shown in Table 7.

Analog Output Calibration Registers	Sampling Frequency
0	> 54 kHz and < = 108 kHz
1	> 30 kHz and < = 54 kHz
2	> 108 kHz and < = 216 kHz

Table 7: Analog Output Calibration Registers

Each potentiometer has three values: a "factory" calibration value that is programmed at the factory and represents a known accurate configuration, a "user" calibration value that you can modify, and a "wiper" value, which is the current value. In most cases, the wiper and user calibration values are the same. They differ only during the actual calibration process when the wiper value is modified until the desired value is reached.

This section describes how to calibrate the analog output circuitry.

Opening the Subsystem

Open the file corresponding to the analog input subsystem (/dev/dt7837-aout) using the Linux command **open**, described on page 22.

Calibrating the Offset

To calibrate the offset of an analog output channel, connect a digital multimeter (DMM) to the specified analog output channel. Then, use the **ioctl - IOCTL_OFFSET_POT_SET** command to specify the analog output channel that you want to calibrate, the value to write to the potentiometer, whether the calibration type is a user (0) or factory (1) calibration, and the calibration register to update (0 to 2) based on the output frequency used.

You can continue to adjust the offset values by calling this command with different values. Adjust the potentiometer until you measure a value as close to 0 V as possible. The offset calibration register is automatically written to the wiper of the associated potentiometer.

When you are satisfied that the offset calibration value is correct, call the **ioctl** - **IOCTL_OFFSET_POT_WIPER_SET** command. You specify the analog input channel (potentiometer) that you want to calibrate and the value to write to the potentiometer. The value is written to a calibration register in non-volatile EEPROM.

These steps should be repeated until all the offset calibration registers (for all output frequency ranges) are calibrated for each analog output channel.

Note that you can read the value of the offset potentiometers using the **ioctl** - **IOCTL_OFFSET_POT_GET** and **ioctl** - **IOCTL_OFFSET_POT_WIPER_GET** commands.

Calibrating the Gain

To calibrate the gain of an analog output channel, connect a DMM to the specified analog output channel.

Then, use the **ioctl - IOCTL_GAIN_POT_SET** command to specify the analog output channel that you want to calibrate, the value to write to the potentiometer, whether the calibration type is a user or factory calibration, and the calibration register to update (0 to 2) based on the output frequency used.

Note: This command will block all other operations for at least 5 ms.

You can continue to adjust the gain calibration values by calling this command with different values. Adjust the potentiometer until you measure a value as close to +9.375 V as possible. The gain calibration register is automatically written to the wiper of the associated potentiometer.

When you are satisfied that the gain calibration value is correct, call the **ioctl -IOCTL_GAIN_POT_WIPER_SET** command. You specify the analog input channel (potentiometer) that you want to calibrate and the value to write to the potentiometer. The value is written to a calibration register in non-volatile EEPROM.

These steps should be repeated until all the gain calibration registers (for all output frequencies) are calibrated for each analog output channel.

Note that you can read the value of the gain potentiometers using the **ioctl** - **IOCTL_GAIN_POT_GET** and **ioctl** - **IOCTL_GAIN_POT_WIPER_GET** commands.

Closing the Subsystem

When finished, close the analog output subsystem using the **close** command, described on page 23.

Modifying the State of the User LEDs on the Module

The DT7837 has eight user LEDs, shown in Figure 5, on the ARM block (bottom board). Header J10, also shown in Figure 5, provides debug pins that correspond to the user LEDs. The value of a debug pin reflects the state of the corresponding user LED, where the pin has a value of 0 if the LED is off or a value of 1 if the LED is on.



Figure 5: User LEDs on the DT7837 Module

Table 8 lists the pin descriptions of header J10 on the DT7837.

Pin	Pin Description
1	Debug_D7; corresponds to user LED 7.
2	Debug_D6; corresponds to user LED 6.
3	Debug_D5; corresponds to user LED 5.
4	Debug_D4; corresponds to user LED 4.
5	Debug_D3; corresponds to user LED 3.
6	Debug_D2; corresponds to user LED 2.
7	Debug_D1; corresponds to user LED 1.
8	Debug_D0; corresponds to user LED 0.
9	Digital Ground
10	Digital Ground

Table 8: Debug Pins of Header J10

You can turn the LEDs either on or off using the **ioctl - IOCTL_LED_SET** command. You can also return the state of the user LEDs using the **ioctl - IOCTL_LED_GET** command.

Note that to execute these commands, you must first open a file corresponding to any of the supported subsystems of the DT7837 (/dev/dt7837-ain, dev/dt7837-din, /dev/dt7837-dout, /dev/dt7837-tach, dev/dt7837-measure, /dev/dt7837-ctr-tmr, or /dev/dt7837_stream_in) using the Linux command **open**, described on page 22. When finished, close the file using the **close**, described on page 23.

Sending Data to or Receiving Data from the Host USB Application

The kernel module for the DT7837 USB gadget driver is a simplified version of gadgetfs and has the same kind of functionality. When loaded, it enumerates as a USB gadget, creates a specified number of bulk USB endpoints, and exposes each end point as a virtual file in the **/dev** directory.

A user mode application can perform file operations on each virtual file to send data to the USB host through an IN endpoint and receive data from the USB host through an OUT endpoint. The module serves as a conduit for data and does not care about the data itself.

Note that the directions are relative to the USB host. Therefore, to send data to the host USB, you must write to an IN endpoint. To receive data from a USB host, you must read data from an OUT endpoint.

This section describes file I/O operations used by the USB gadget driver to send data to and receive data from the USB host.

Opening the File

Open the file corresponding to the endpoint that you want to open using the Linux command **open**, described on page 22.

The files are named as follows:

- /dev/dt7837-ep1in
- /dev/dt7837-ep1out
- /dev/dt7837-ep2in
- /dev/dt7837-ep2out
- /dev/dt7837-ep3out
- /dev/dt7837-ep4out
- /dev/dt7837-ep5out

A file can be opened by only one reader/writer at a time. However, different files can be opened simultaneously. Once a file is closed, it can be re-opened.

Sending Data to the USB Host

To send data to the USB host, you must write to file corresponding to an IN endpoint.

If you want to perform a blocking write, use the Linux command **write**. Because it is blocking, no other operation can be performed until the **write** operation is complete.

If you want to perform a non-blocking, asynchronous write operation, use the Linux command **aio_write**. Because it is non-blocking, other operations can be performed while the **aio_write** operation is in progress.

Receiving Data from the USB Host

To receive data from a USB host, you must read data from a file corresponding to an OUT endpoint.

If you want to perform a blocking read operation, use the Linux command **read**. Because it is blocking, no other operation can be performed until the **read** operation is complete.

If you want to perform a non-blocking, asynchronous read operation, use the Linux command **aio_read**. Because it is non-blocking, other operations can be performed while the **aio_read** operation is in progress.

Eliminating Data from the Endpoint

To eliminate the data from either an IN or OUT endpoint, use the Linux command **flush**.

Closing the File

When you are finished using the file, close the file using the **close** command, described on page 23.



File I/O Command Reference

close
ioctl - IOCTL_ACQ_STATUS_GET
ioctl - IOCTL_AIN_CNF_GET
ioctl - IOCTL_AIN_CNF_SET69
ioctl - IOCTL_ARM_SUBSYS
ioctl - IOCTL_CHAN_MASK_GET
ioctl - IOCTL_CHAN_MASK_SET75
ioctl - IOCTL_CT_CFG_GET
ioctl - IOCTL_CT_CFG_SET
ioctl - IOCTL_GAIN_POT_GET92
ioctl - IOCTL_GAIN_POT_SET94
ioctl - IOCTL_GAIN_POT_WIPER_GET
ioctl - IOCTL_GAIN_POT_WIPER_SET99
ioctl - IOCTL_LED_GET 101
ioctl - IOCTL_LED_SET
ioctl - IOCTL_MCTR_CFG_GET 105
ioctl - IOCTL_MCTR_CFG_SET 111
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET123
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET123ioctl - IOCTL_SAMPLE_CLK_GET125
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET123ioctl - IOCTL_SAMPLE_CLK_GET125ioctl - IOCTL_SAMPLE_CLK_SET127
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET123ioctl - IOCTL_SAMPLE_CLK_GET125ioctl - IOCTL_SAMPLE_CLK_SET127ioctl - IOCTL_START_SUBSYS.129
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET123ioctl - IOCTL_SAMPLE_CLK_GET125ioctl - IOCTL_SAMPLE_CLK_SET127ioctl - IOCTL_START_SUBSYS129ioctl - IOCTL_START_TRIG_CNF_GET131
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET.123ioctl - IOCTL_SAMPLE_CLK_GET125ioctl - IOCTL_SAMPLE_CLK_SET127ioctl - IOCTL_START_SUBSYS.129ioctl - IOCTL_START_TRIG_CNF_GET131ioctl - IOCTL_START_TRIG_CNF_SET135
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET.123ioctl - IOCTL_SAMPLE_CLK_GET.125ioctl - IOCTL_SAMPLE_CLK_SET127ioctl - IOCTL_START_SUBSYS.129ioctl - IOCTL_START_TRIG_CNF_GET131ioctl - IOCTL_START_TRIG_CNF_SET135ioctl - IOCTL_STOP_SUBSYS.139
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET.123ioctl - IOCTL_SAMPLE_CLK_GET125ioctl - IOCTL_SAMPLE_CLK_SET127ioctl - IOCTL_START_SUBSYS.129ioctl - IOCTL_START_TRIG_CNF_GET131ioctl - IOCTL_START_TRIG_CNF_SET135ioctl - IOCTL_STOP_SUBSYS.139ioctl - IOCTL_TACH_CFG_GET140
ioctl - IOCTL_MCTR_CFG_SET.111ioctl - IOCTL_OFFSET_POT_GET.117ioctl - IOCTL_OFFSET_POT_SET119ioctl - IOCTL_OFFSET_POT_WIPER_GET121ioctl - IOCTL_OFFSET_POT_WIPER_SET.123ioctl - IOCTL_SAMPLE_CLK_GET125ioctl - IOCTL_SAMPLE_CLK_SET127ioctl - IOCTL_START_SUBSYS.129ioctl - IOCTL_START_TRIG_CNF_GET131ioctl - IOCTL_START_TRIG_CNF_SET135ioctl - IOCTL_STOP_SUBSYS.139ioctl - IOCTL_TACH_CFG_GET140ioctl - IOCTL_TACH_CFG_SET142

Note: Refer to your Linux documentation for information on all other Linux commands, including **read**, **write**, **aio_read**, **aio_write**, **io_setup**, **io_submit**, **io_getevents**, **io_cancel**, and **io_destroy**.

close

Description	Closes a file descriptor that is associated with a subsystem, stream, or endpoint of the DT7837 module.
Syntax	<pre>int close(int fd);</pre>
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor of the subsystem, stream, or endpoint to close.
Returns	0 = Success; $-1 =$ Failure
Notes	You must open each subsystem, stream, or endpoint that you want to use with the open command, described on page 144, before you can configure its parameters and perform an I/O operation.
Example	The following command closes the file associated with the input stream of the DT7837:
	#define DEV_STREAM_IN "/dev/dt7837-stream-in"
	<pre>close(DEV_STREAM_IN);</pre>
See Also	open, described on page 144.

ioctl - IOCTL_ACQ_STATUS_GET

Description	Returns the status of acquisition or the status of the output operation.
Syntax	<pre>int ioctl(int fd, IOCTL_ACQ_STATUS_GET,</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or the output stream (/dev/dt7837-stream-out).
Name:	pAcq_Status
Data Type:	acq_status_t enumeration
Description:	A pointer to a variable that specifies the status of acquisition (if the file $/dev/dt7837$ -stream-in was specified as the file descriptor) or the status of the output operation (if the file $/dev/dt7837$ -stream-out was specified as the file descriptor).
	<i>The acq_status_t</i> enumeration is described as follows:
	<pre>{ acq_status_armed = 0x01, acq_status_triggered = 0x02, acq_status_fifo_empty = 0x04, acq_status_fifo_full = 0x08, }acq_status_t;</pre>
	where,
	• <i>acq_status_armed</i> specifies that the associated stream is armed and waiting for a trigger (value 0x1).
	• <i>acq_status_triggered</i> specifies that the associated stream has been triggered and is active (value 0x2).
	• <i>acq_status_fifo_empty</i> indicates that the FIFO is empty (no data) for the associated stream (value 0x4).
	 acq_status_fifo_full indicates that the FIFO is full of data for the associated stream.
Returns	0 = Success; < 0 = Failure
Example	The following example returns the status of acquisition for the input stream:
	<pre>acq_status_t Acq_Status; if (ioctl(fd_instream, IOCTL_ACQ_STATUS_GET, &Acq_Status)) return (printf_errno_html(conn, errno, "IOCTL_ACQ_STATUS_GET"));</pre>

ioctl - IOCTL_AIN_CNF_GET

Description	n Returns the configuration of the analog input subsystem.		
Syntax	<pre>int ioctl(int fd, IOCTL_AIN_CFG_GET, struct dt78xx_ain_config_t *pAIN_config);</pre>		
Include File	DT78XX_IOCTL.H		
Arguments			
Name:	fd		
Data Type:	int		
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain).		
Name:	pAIN_config		
Data Type:	dt78xx_ain_config_t structure		
Description:	A pointer to a structure that defines the configuration of the analog input subsystem.		
	<i>dt78xx_ain_config_t</i> is defined as follows:		
	<pre>typedef structattribute ((packed)) { uint16_t ain; uint16_t gain; uint16_t ac_coupling; uint16_t current_on; uint16_t differential; uint16_t unused; } dt78xx_ain_config_t;</pre>		
Structure Element Name:	ain		
Data Type:	uint16_t		
Description:	Specifies which analog input channel to configure, where 0 represents analog input channel 0, 1 represents analog input channel 1, 2 represents analog input channel 2, and 3 represents analog input channel 3.		
Structure Element Name:	gain		
Data Type:	uint16_t		
Description:	Specifies the gain to use for the specified analog input channel. Values are 1 for a gain of 1 or 10 for a gain of 10.		
Structure Element Name:	ac_coupling		
Data Type:	uint16_t		
Description:	Specifies the coupling type for the specified analog input channel. Values are 0 for dc coupling or 1 for ac coupling.		

Structure Element Name:	current_on			
Data Type:	uint16_t			
Description:	Specifies whether the 4 mA current source is on or off for the specified analog input channel. If this value is 0, the current source is off. If this value is 1, the current source is on.			
Structure Element Name:	differential			
Data Type:	uint16_t			
Description:	Specifies whether the specified analog input channel is a differential input or a single-ended input. If this value is 0, the channel is configured for a single-ended input. If this value is 1, the channel is configured for a differential input. The DT7837 supports single-ended inputs only.			
Structure Element Name:	unused			
Data Type:	uint16_t			
Description:	Reserved for future use.			
Returns	0 = Success; $< 0 =$ Failure			
Example	The following example returns the configuration of the analog input subsystem:			
	<pre>dt78xx_ain_config_t AIN_cfg; if (ioctl(fd_ain, IOCTL_AIN_CFG_GET, &AIN_cfg)) { perror("IOCTL_AIN_CFG_GET"); goto finish; }</pre>			
See Also	IOCTL_AIN_CFG_SET, described on page 69.			

ioctl - IOCTL_AIN_CNF_SET

Description	Specifies the configuration of the analog input subsystem.				
Syntax	<pre>int ioctl(int fd, IOCTL_AIN_CFG_SET, struct dt78xx_ain_config_t *pAIN_config);</pre>				
Include File	DT78XX_IOCTL.H				
Arguments					
Name:	fd				
Data Type:	int				
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain).				
Name:	pAIN_config				
Data Type:	dt78xx_ain_config_t structure				
Description:	A pointer to a structure that defines the configuration of the analog input subsystem.				
	<i>dt78xx_ain_config_t</i> is defined as follows:				
	<pre>typedef structattribute ((packed)) { uint16_t ain; uint16_t gain; uint16_t ac_coupling; uint16_t current_on; uint16_t differential; uint16_t unused; } dt78xx_ain_config_t;</pre>				
Structure Element Name:	ain				
Data Type:	uint16_t				
Description:	Specifies which analog input channel to configure, where 0 represents analog input channel 0, 1 represents analog input channel 1, 2 represents analog input channel 2, and 3 represents analog input channel 3.				
Structure Element Name:	gain				
Data Type:	uint16_t				
Description:	Specifies the gain to use for the specified analog input channel. Values are 1 for a gain of 1 or 10 for a gain of 10.				
Structure Element Name:	ac_coupling				
Data Type:	uint16_t				
Description:	Specifies the coupling type for the specified analog input channel. Values are 0 for dc coupling or 1 for ac coupling.				

Structure Element Name:	current_on
Data Type:	uint16_t
Description:	Specifies whether the 4 mA current source is on or off for the specified analog input channel. If this value is 0, the current source is off. If this value is 1, the current source is on.
Structure Element Name:	differential
Data Type:	uint16_t
Description:	Specifies whether the specified analog input channel is a differential input or a single-ended input. If this value is 0, the channel is configured for a single-ended input. If this value is 1, the channel is configured for a differential input. The DT7837 supports single-ended inputs only.
Structure Element Name:	unused
Data Type:	uint16_t
Description:	Reserved for future use.
Returns	0 = Success; $< 0 =$ Failure
Example	The following example specifies the configuration of the analog input subsystem. In this case, analog input channel 0 is configured with a gain of 1, dc coupling, and with the 4 mA current source off.
	<pre>#define DEV_AIN "/dev/dt7837-ain" fd_ain = open(DEV_AIN, O_RDWR);</pre>
	<pre>dt78xx_ain_config_t AIN_cfg; AIN_cfg.ain = 0; AIN_cfg.gain = 1; //x1 gain AIN_cfg.ac_coupling = 0; //dc coupling AIN_cfg.current_on = 0; //current source off if (ioctl(fd_ain, IOCTL_AIN_CFG_SET, &AIN_cfg)) { fprintf(stderr, "IOCTL_AIN_CFG_SET ERROR</pre>
	goto finish;
See Also	IOCTL_AIN_CFG_GET, described on page 67.

ioctl - IOCTL_ARM_SUBSYS

Description	Arms the input or output stream to detect the trigger that will start the continuous operation.			
Syntax	<pre>int ioctl(int fd, IOCTL_ARM_SUBSYS, int unused);</pre>			
Include File	DT78XX_IOCTL.H			
Arguments				
Name:	fd			
Data Type:	int			
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or output stream (/dev/dt7837-stream-out).			
Name:	unused			
Data Type:	int			
Description:	Not used; reserved for future use.			
Returns	0 = Success; $< 0 = $ Failure			
Example	The following example arms the input stream, <i>fd_instream</i> , that was previously opened:			
	<pre>tmp = 0; if (ioctl(fd_in_stream, IOCTL_ARM_SUBSYS, &tmp)) { err = errno; return (printf_errno(conn, err, "IOCTL_ARM_SUBSYS")); }</pre>			

See Also

ioctl - IOCTL_START_SUBSYS, described on page 129.

ioctl - IOCTL_CHAN_MASK_GET

Description	Returns the channels that are enabled for acquisition in the input stream or that are enabled for output in the output stream.	
Syntax	<pre>int ioctl(int fd, IOCTL_CHAN_MASK_GET,</pre>	
Include File	DT78XX_IOCTL.H	
Arguments		
Name:	fd	
Data Type:	int	
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or output stream (/dev/dt7837-stream-out).	
Name:	*pChannels	
Data Type:	chan_mask_t enumeration	
Description:	A pointer to a 32-bit value that specifies the channels that are enabled for acquisition in the input stream or enabled for output in the output stream. The <i>channel mask t</i> enumeration, described below, is used to specify the	

values associated with the channels in the input stream and output stream:

typedef enum {

	chan_mask_dout7	=	(1<<31),
	chan_mask_dout6	=	(1<<30),
	chan_mask_dout5	=	(1<<29),
	chan_mask_dout4	=	(1<<28),
	chan_mask_dout3	=	(1<<27),
	chan_mask_dout2	=	(1<<26),
	chan_mask_dout1	=	(1<<25),
	chan_mask_dout0	=	(1<<24),
	chan_mask_aout3	=	(1<<19)
	chan_mask_aout2	=	(1<<18),
	chan_mask_aout1	=	(1<<17),
	chan_mask_aout0	=	(1<<16),
	chan_mask_din	=	(1<<11),
	chan_mask_meas_ctr	=	(1<<10),
	chan_mask_ctr_timer	=	(1<<9),
	chan_mask_tach	=	(1<<8),
	chan_mask_ain7	=	(1<<7),
	chan_mask_ain6	=	(1<<6),
	chan_mask_ain5	=	(1<<5),
	chan_mask_ain4	=	(1<<4),
	chan_mask_ain3	=	(1<<3),
	chan_mask_ain2	=	(1<<2),
	chan_mask_ain1	=	(1<<1),
	chan_mask_ain0	=	(1<<0),
}	chan_mask_t		
Description (cont.): where,

- *chan_mask_dout7 is* the line 7 of the digital output port (bit 31 value 0x8000000).
- *chan_mask_dout6 is* the line 6 of the digital output port (bit 30 value 0x4000000).
- *chan_mask_dout5 is* the line 5 of the digital output port (bit 29 value 0x2000000).
- *chan_mask_dout4 is* the line4 of the digital output port (bit 28 value 0x1000000).
- *chan_mask_dout3 is* the line3 of the digital output port (bit 27 value 0x8000000).
- *chan_mask_dout2 is* the line1 of the digital output port (bit 26 value 0x4000000).
- *chan_mask_dout1 is* the line1 of the digital output port (bit 25 value 0x2000000).
- *chan_mask_dout0 is* the line 0 of the digital output port (bit 24 value 0x1000000).
- *chan_mask_aout3* is analog output 3 (bit 19 value 0x80000). Note that the DT7837 does not support this channel.
- *chan_mask_aout2* is analog output 2 (bit 18 value 0x40000). Note that the DT7837 does not support this channel.
- *chan_mask_aout1* is analog output 1 (bit 17 value 0x20000). Note that the DT7837 does not support this channel.
- *chan_mask_aout0* is analog output 0 (bit 16 value 0x10000).
- *chan_mask_din is* the digital input port (bit 11 value 0x800).
- *chan_mask_meas_ctr* is the measure counter (bit 10 value 0x400).
- *chan_mask_tach* is the tachometer input (bit 8 value 0x100).
- *chan_mask_ain7* is analog input channel 7 (bit 7 value 0x80). Note that the DT7837 does not support this channel.
- *chan_mask_ain6* is analog input channel 6 (bit 6 value 0x40). Note that the DT7837 does not support this channel.
- *chan_mask_ain5* is analog input channel 5 (bit 5 value 0x20). Note that the DT7837 does not support this channel.
- *chan_mask_ain4* is analog input channel 4 (bit 4 value 0x10). Note that the DT7837 does not support this channel.
- *chan_mask_ain3* is analog input channel 3 (bit 3 value 0x8).
- *chan_mask_ain2* is analog input channel 2 (bit 2 value 0x4).
- *chan_mask_ain1* is analog input channel 1 (bit 1 value 0x2).
- *chan_mask_ain0* is analog input channel 0 (bit 0 value 0x1).

Returns	0 = Success; $< 0 =$ Failure
Example	The following example returns the channels that were enabled in the input stream, <i>fd_instream</i> :
	<pre>ret = ioctl(fd_instream, IOCTL_CHAN_MASK_GET, &chan_mask);</pre>
See Also	IOCTL_CHAN_MASK_SET, described on page 75.

ioctl - IOCTL_CHAN_MASK_SET

Description	Specifies the channels to enable for acquisition in the input stream or enabled for output in the output stream.
Syntax	<pre>int ioctl(int fd, IOCTL_CHAN_MASK_SET,</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or output stream (/dev/dt7837-stream-out).
Name:	*pChannels
Data Type:	chan_mask_t enumeration
Description:	A pointer to a 32-bit value that specifies the channels that are enabled for acquisition in the input stream or enabled for output in the output stream. The <i>channel_mask_t</i> enumeration, described below, is used to specify the

values associated with the channels in the input stream and output stream:

typedef enum { chan_mask_dout7 = (1 < < 31),chan_mask_dout6 = (1 < < 30),chan_mask_dout5 = (1<<29), chan_mask_dout4 = (1 < < 28),chan_mask_dout3 = (1<<27), chan_mask_dout2 = (1 < < 26),chan_mask_dout1 = (1 < < 25),chan_mask_dout0 = (1 < < 24),= (1 < < 19)chan_mask_aout3 chan_mask_aout2 = (1 < < 18),chan_mask_aout1 = (1<<17), chan_mask_aout0 = (1<<16), chan_mask_din = (1<<11), = (1<<10), chan_mask_meas_ctr chan_mask_ctr_timer = (1<<9),</pre> chan_mask_tach = (1<<8), chan_mask_ain7 = (1<<7), chan_mask_ain6 = (1<<6), chan_mask_ain5 = (1<<5), chan_mask_ain4 = (1<<4), chan_mask_ain3 = (1<<3), chan_mask_ain2 = (1<<2), chan_mask_ain1 = (1<<1), chan_mask_ain0 = (1<<0), } chan_mask_t

Description (cont.): where,

- *chan_mask_dout7 is* the line 7 of the digital output port (bit 31 value 0x8000000).
- *chan_mask_dout6 is* the line 6 of the digital output port (bit 30 value 0x4000000).
- *chan_mask_dout5 is* the line 5 of the digital output port (bit 29 value 0x2000000).
- *chan_mask_dout4 is* the line4 of the digital output port (bit 28 value 0x1000000).
- *chan_mask_dout3 is* the line3 of the digital output port (bit 27 value 0x8000000).
- *chan_mask_dout2 is* the line1 of the digital output port (bit 26 value 0x4000000).
- *chan_mask_dout1 is* the line1 of the digital output port (bit 25 value 0x200000).
- *chan_mask_dout0 is* the line 0 of the digital output port (bit 24 value 0x1000000).
- *chan_mask_aout3* is analog output 3 (bit 19 value 0x80000). Note that the DT7837 does not support this channel.
- *chan_mask_aout2* is analog output 2 (bit 18 value 0x40000). Note that the DT7837 does not support this channel.
- *chan_mask_aout1* is analog output 1 (bit 17 value 0x20000). Note that the DT7837 does not support this channel.
- *chan_mask_aout0* is analog output 0 (bit 16 value 0x10000).
- *chan_mask_din is* the digital input port (bit 11 value 0x800).
- *chan_mask_meas_ctr* is the measure counter (bit 10 value 0x400).
- *chan_mask_tach* is the tachometer input (bit 8 value 0x100).
- *chan_mask_ain7* is analog input channel 7 (bit 7 value 0x80). Note that the DT7837 does not support this channel.
- *chan_mask_ain6* is analog input channel 6 (bit 6 value 0x40). Note that the DT7837 does not support this channel.
- *chan_mask_ain5* is analog input channel 5 (bit 5 value 0x20). Note that the DT7837 does not support this channel.
- *chan_mask_ain4* is analog input channel 4 (bit 4 value 0x10). Note that the DT7837 does not support this channel.
- *chan_mask_ain3* is analog input channel 3 (bit 3 value 0x8).
- *chan_mask_ain2* is analog input channel 2 (bit 2 value 0x4).
- *chan_mask_ain1* is analog input channel 1 (bit 1 value 0x2).
- *chan_mask_ain0* is analog input channel 0 (bit 0 value 0x1).

Returns 0 = Success; < 0 = Failure
Example The following example specifies the channels to enable in the input stream.
In this example, analog input channels 0 and 2 are enabled in the input
stream; this equivalent to value 0x5:
 chan_mask_t chan_mask = chan_mask_ain0|chan_mask_ain2;
 if (ioctl(fd_stream, IOCTL_CHAN_MASK_SET, &chan_mask))
 {
 fprintf(stderr, "IOCTL_CHAN_MASK_SET ERROR
 %d \"%s\"\n", errno, strerror(errno));
 goto _cleanup;</pre>

See Also IOCTL_CHAN_MASK_GET, described on page 72.

ioctl - IOCTL_CT_CFG_GET

Description	Returns the configuration of the counter/timer channel.	
Syntax	<pre>int ioctl(int fd, IOCTL_CT_CFG_GET, struct dt78xx_ct_config_t *pCT_config);</pre>	
Include File	DT78XX_IOCTL.H	
Arguments		
Name:	fd	
Data Type:	int	
Description:	The file descriptor associated with the counter/timer subsystem (/dev/dt7837-ctr-tmr).	
Name:	pCT_config	
Data Type:	dt78xx_ct_config_t structure	
Description:	A pointer to a structure that defines the configuration of the counter/timer channel.	
	<i>dt78xx_ct_config_t</i> is defined as follows:	
	<pre>typedef structattribute ((packed)) { ct_mode_t mode; ct_gate_t gate; uint8_t ext_gate_din; uint8_t ext_clk; uint8_t ext_clk_din; uint8_t unused; union { structattribute ((packed)) { uint32_t period; uint8_t out; uint8_t out; uint8_t out_hi; }divider; </pre>	
	<pre>structattribute ((packed)) { uint32_t period; uint32_t pulse; uint8_t out; uint8_t out_hi; uint8_t retriggerable; }one_shot; }; }dt78xx ct config t:</pre>	

```
Structure Element Name: mode
Data Type: ct_mode_t
Description: This enumeration defines the operation mode of the counter/timer
channel. It is defined as follows:
    typedef enum
    {
        ct_mode_idle,
        ct_mode_lshot,
        ct_mode_lshot,
        ct_mode_counter,
        ct_mode_measure,
        }ct_mode_t;
```

- *ct_mode_idle* specifies idle mode. If you specify this mode, the counter no longer drives the clock output signal that is assigned to one of the general-purpose output signal (pins 11 to 18) of the Digital connector.
- ct_mode_1shot specifies one-shot mode. If you specify this mode, values for the one_shot structure specified in dt78xx_ct_config_t are required. Specify one-shot mode if you want to generate non-retriggerable one-shot pulses.
- ct_mode_divider specifies rate generation mode. If you specify this mode, values for the *divider* structure specified in *dt78xx_ct_config_t* are required. In rate generation mode, a continuous, active high, pulse output signal is generated from the counter's output signal. You can use this pulse output signal as an external clock to pace other operations, such as analog input, analog output, or other counter/timer operations. The pulse output operation starts whenever the counter's gate signal is at the specified level.
- *ct_mode_counter* specifies event counting mode. In event counting mode, the number of rising edges that occur on the counter's clock input signal are counted when the counter's gate signal is active. The module can count a maximum of 4,294,967,296 events before the counter rolls over to 0 and starts counting again.
- *ct_mode_measure* is currently not supported.

Structure Element Name:	gate
Data Type:	ct_gate_t enumeration
Description:	Defines the gate type to use for the specified counter/timer operation mode. It is defined as follows:
	typedef enum { ct gate none = 0x1.

```
ct_gate_none = 0x1,
ct_gate_ext_hi = 0x2,
ct_gate_ext_lo = 0x3
}ct_gate_t;
```

- ct_gate_none specifies a software gate type. With this gate type, the ioctl

 IOCTL_START_SUBSYS command, described on page 129, starts the counter/timer operation immediately after execution. (No general-purpose input signal is required if this gate type is selected.) To stop the operation, use the ioctl IOCTL_START_SUBSYS command, described on page 139.
- ct_gate_ext_hi specifies a logic high or rising edge gate type. For event counting (counter) and rate generation (divider) mode, the operation is enabled when the counter's gate signal is high and is disabled when the counter's gate signal is low. For one-shot and repetitive one-shot (1shot) mode, the operation is enabled when the counter's gate signal goes from a low to a high transition and is disabled when the counter's gate signal goes from a high to a low transition.
- ct_gate_ext_lo specifies a logic low or falling edge gate type. For event counting (counter) and rate generation (divider) mode, the operation is enabled when the counter's gate signal is low and is disabled when the counter's gate signal is high. For one-shot and repetitive one-shot mode (1shot) mode, the operation is enabled when the counter's gate signal goes from a high to a low transition and is disabled when the counter's gate signal goes from a low to a high transition.

Structure Element Name:	ext_gate_din
Data Type:	uint8_t
Description:	Specifies which general-purpose input signal (pins 1 to 8) on the Digital connector to use for the external C/T gate input signal.
	Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
	Pin 2 corresponds to bit 1 of the digital input port (value 0x2).
	Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
	Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
	Pin 5 corresponds to bit 4 of the digital input port (value 0x10).
	Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
	Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
	Pin 8 corresponds to bit 7 of the digital input port (value 0x80).
Structure Element Name:	ext_clk
Data Type:	uint8_t
Description:	Specifies whether to use the internal C/T clock or an external C/T clock. When this value is 0, the internal C/T clock is used. When this value is 1, an external C/T clock is used.
	Note that in one-shot mode, the internal C/T clock is more useful than an external C/T clock. However, in event counting mode, up/down counting, and measure mode, an external C/T clock source is generally used.
Structure Element Name:	ext_clk_din
Data Type:	uint8_t
Description:	Specifies which general-purpose input signal (pins 1 to 8) on the Digital connector to use for the external C/T clock input signal.
	Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
	Pin 2 corresponds to bit 1 of the digital input port (value 0x2).
	Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
	Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
	Pin 5 corresponds to bit 4 of the digital input port (value $0x10$).
	Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
	Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
	Pin 8 corresponds to bit 7 of the digital input port (value 0x80).
Structure Element Name:	unused
Data Type:	uint8_t
Description:	This variable is reserved for future use.

Structure Element Name:	period
Data Type:	uint32_t
Description:	Used by the <i>divider</i> and <i>one_shot</i> structures, specifies the number of input clock cycles used to create one period of the C/T output signal.
	The period of the output pulse is determined by the C/T clock source (<i>ext_clk</i> element of the <i>dt78xx_ct_config_t</i> structure, described on page 81.) You can output pulses using a maximum frequency of 24 MHz (if using the internal C/T clock) or 5 MHz (if using the external C/T clock). Note, however, that the integrity of the signal degrades at frequencies greater than 10 MHz.
	For example, if you are using an external C/T clock of10000 Hz as the input clock source of the counter/timer, and you want to generate a output signal of 1000 Hz with a 20% duty cycle, specify a period of 10 (10000 Hz divided by 10 is 1000 Hz) and a pulse width of 2 (the period multiplied by 20%). This is illustrated in Figure 6.

Note that the polarity of the output pulse can be programmed using the *out_hi* element of the *divider* or *one_shot* structure.



Figure 6: Example of a Pulse Width and Period

Structure Element Name:	pulse
Data Type:	uint32_t
Description:	Used by the <i>divider</i> and <i>one_shot</i> structures, specifies the number of input clock cycles used to create the active pulse width (or duty cycle) of the C/T output signal. This value must be less than the <i>period</i> . See Figure 6 for an example.

Structure Element Name:	out
Data Type:	uint8_t
Description:	Used by the <i>divider</i> and <i>one_shot</i> structures, specifies which general-purpose output signal (pins 11 to 18) on the Digital connector to use for the external C/T clock output signal.
	Pin 11 corresponds to bit 0 of the digital output port (value 0x1).
	Pin 12 corresponds to bit 1 of the digital output port (value 0x2).
	Pin 13 corresponds to bit 2 of the digital output port (value 0x4).
	Pin 14 corresponds to bit 3 of the digital output port (value 0x8).
	Pin 15 corresponds to bit 4 of the digital output port (value $0x10$).
	Pin 16 corresponds to bit 5 of the digital output port (value 0x20).
	Pin 17 corresponds to bit 6 of the digital output port (value 0x40).
	Pin 18 corresponds to bit 7 of the digital output port (value 0x80).
Structure Element Name:	out_hi
Data Type:	uint8_t
Description:	Used by the <i>divider</i> and <i>one_shot</i> structures, specifies whether the output pulse for a rate generation or non-retriggerable one-shot operation is active high (1) or active low (0).
Structure Element Name:	retriggerable
Data Type:	uint8_t
Description:	Used by the <i>one-shot</i> structure, specifies whether or not to repeat the one-shot output pulse.
	Currently, only a value of 0 is supported.
	When <i>retriggerable</i> is 0, the counter generates a single pulse output signal whenever it detects the active gate signal (after the pulse period from the previous output pulse expires). Any gate signals that occur while the pulse is being output are not detected by the module. The module continues to output pulses until you stop the operation with ioctl - IOCTL_STOP_SUBSYS , described on page 139. You can use this mode to clean up a poor clock input signal by changing its pulse width, and then outputting it.
Returns	0 = Success; $< 0 = $ Failure
Notes	The value of the counter clock output signal (assigned to one of the general-purpose output signals) can be overwritten by writing to the digital output subsystem. Refer to page 53 for more information.
	If you assigned a general-purpose input signal as a counter clock or gate input (or external trigger), you can read the value of the signal as you would any other digital input signal. Refer to page 52 for more information on reading digital input values.

Example	The following example returns the configuration of the counter/timer channel:	
	<pre>if (ioctl(ct_file, IOCTL_CT_CFG_GET, &ct_cfg)) {</pre>	
	<pre>perror("IOCTL_CT_CFG_GET"); goto finish;</pre>	
	}	
0		

See Also IOCTL_CT_CFG_SET, described on page 85.

ioctl - IOCTL_CT_CFG_SET

Description	Specifies the configuration of the counter/timer channel.
Syntax	<pre>int ioctl(int fd, IOCTL_CT_CFG_SET, struct dt78xx_ct_config_t *pCT_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the counter/timer subsystem (/dev/dt7837-ctr-tmr).
Name:	pCT_config
Data Type:	dt78xx_ct_config_t structure
Description:	A pointer to a structure that defines the configuration of the counter/timer channel.
	<i>dt78xx_ct_config_t</i> is defined as follows:
	<pre>{ ct_mode_t mode; ct_gate_t gate; uint8_t ext_gate_din; uint8_t ext_clk; uint8_t ext_clk_din; uint8_t unused; union { structattribute ((packed)) { uint32_t pulse; uint8_t out; uint8_t out; uint8_t out; uint8_t out_hi; }divider; } } </pre>
	<pre>structattribute ((packed)) { uint32_t period; uint32_t pulse; uint8_t out; uint8_t out_hi; uint8_t retriggerable; }one_shot; }; }dt78xx ct config t;</pre>

```
Structure Element Name: mode
Data Type: ct_mode_t
Description: This enumeration defines the operation mode of the counter/timer
channel. It is defined as follows:
    typedef enum
    {
        ct_mode_idle,
        ct_mode_lshot,
        ct_mode_livider,
        ct_mode_counter,
        ct_mode_measure,
```

}ct_mode_t;

- ct_mode_idle specifies idle mode. If you specify this mode, the counter no longer drives the clock output signal that is assigned to one of the general-purpose output signals (pins 11 to 18) of the Digital connector.
- ct_mode_1shot specifies one-shot mode. If you specify this mode, values for the one_shot structure specified in dt78xx_ct_config_t are required. Specify one-shot mode if you want to generate non-retriggerable one-shot pulses.
- ct_mode_divider specifies rate generation mode. If you specify this mode, values for the *divider* structure specified in *dt78xx_ct_config_t* are required. In rate generation mode, a continuous pulse output signal is generated from the counter's output signal. You can use this pulse output signal as an external clock to pace other operations, such as analog input, analog output, or other counter/timer operations. The pulse output operation starts whenever the counter's gate signal is at the specified level.
- *ct_mode_counter* specifies event counting mode. In event counting mode, the number of rising edges that occur on the counter's clock input signal are counted when the counter's gate signal is active. The module can count a maximum of 4,294,967,296 events before the counter rolls over to 0 and starts counting again.
- *ct_mode_measure* is currently not supported.

Structure Element Name: gate

Data Type: ct_gate_t enumeration

Description: Defines the gate type to use for the specified counter/timer operation mode. It is defined as follows:

```
typedef enum
{
   ct_gate_none = 0x1,
   ct_gate_ext_hi = 0x2,
   ct_gate_ext_lo = 0x3
}ct_gate_t;
```

- ct_gate_none specifies a software gate type. With this gate type, the ioctl

 IOCTL_START_SUBSYS command, described on page 129, starts the counter/timer operation immediately after execution. (No general-purpose input signal is required if this gate type is selected.) To stop the operation, use the ioctl IOCTL_START_SUBSYS command, described on page 139.
- ct_gate_ext_hi specifies a logic high or rising edge gate type. For event counting (counter) and rate generation (divider) mode, the operation is enabled when the counter's gate signal is high and is disabled when the counter's gate signal is low. For one-shot and repetitive one-shot (1shot) mode, the operation is enabled when the counter's gate signal goes from a low to a high transition and is disabled when the counter's gate signal goes from a high to a low transition.
- ct_gate_ext_lo specifies a logic low or falling edge gate type. For event counting (counter) and rate generation (divider) mode, the operation is enabled when the counter's gate signal is low and is disabled when the counter's gate signal is high. For one-shot and repetitive one-shot (1shot) mode, the operation is enabled when the counter's gate signal goes from a high to a low transition and is disabled when the counter's gate signal goes from a low to a high transition.

Structure Element Name:	ext_gate_din
Data Type:	uint8_t
Description:	Specifies which general-purpose input signal (pins 1 to 8) on the Digital connector to use for the external C/T gate input signal.
	Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
	Pin 2 corresponds to bit 1 of the digital input port (value 0x2).
	Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
	Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
	Pin 5 corresponds to bit 4 of the digital input port (value 0x10).
	Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
	Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
	Pin 8 corresponds to bit 7 of the digital input port (value 0x80).
Structure Element Name:	ext_clk
Data Type:	uint8_t
Description:	Specifies whether to use the internal C/T clock or an external C/T clock. When this value is 0, the internal C/T clock is used. When this value is 1, an external C/T clock is used.
	Note that in one-shot and repetitive one-shot modes (<i>1shot</i>), the internal C/T clock is more useful than an external C/T clock. In event counting (<i>counter</i>) mode, an external C/T clock source is generally used.
	Either clock source can be used for rate generation mode (<i>divider</i>) depending on the application.
Structure Element Name:	ext_clk_din
Data Type:	uint8_t
Description:	Specifies which general-purpose input signal (pins 1 to 8) on the Digital connector to use for the external C/T clock input signal.
	Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
	Pin 2 corresponds to bit 1 of the digital input port (value 0x2).
	Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
	Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
	Pin 5 corresponds to bit 4 of the digital input port (value 0x10).
	Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
	Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
	Pin 8 corresponds to bit 7 of the digital input port (value 0x80).
Structure Element Name:	unused
Data Type:	uint8_t
Description:	This variable is reserved for future use.

Structure Element Name: period Data Type: uint32_t Description: Used by the *divider* and *one_shot* structures, specifies the number of input clock cycles used to create one period of the C/T output signal. The period of the output pulse is determined by the C/T clock source (*ext_clk* element of the *dt78xx_ct_config_t* structure, described on page 88.) You can output pulses using a maximum frequency of 24 MHz (if using the internal C/T clock) or 5 MHz (if using the external C/T clock). Note, however, that the integrity of the signal degrades at frequencies greater than 10 MHz. For example, if you are using an external C/T clock of10000 Hz as the input clock source of the counter/timer, and you want to generate a output signal of 1000 Hz with a 20% duty cycle, specify a period of 10 (10000 Hz divided by 10 is 1000 Hz) and a pulse width of 2 (the period multiplied by 20%). This is illustrated in Figure 7.

Note that the polarity of the output pulse can be programmed using the *out_hi* element of the *divider* or *one_shot* structure.



Figure 7: Example of a Pulse Width and Period

Structure Element Name:	pulse
Data Type:	uint32_t
Description:	Used by the <i>divider</i> and <i>one_shot</i> structures, specifies the number of input clock cycles used to create the active pulse width (or duty cycle) of the C/T output signal. This value must be less than the <i>period</i> . See Figure 6 for an example.

Structure Element Name:	out
Data Type:	uint8_t
Description:	Used by the <i>divider</i> and <i>one_shot</i> structures, specifies which general-purpose output signal (pins 11 to 18) on the Digital connector to use for the external C/T clock output signal.
	Pin 11 corresponds to bit 0 of the digital output port (value 0x1).
	Pin 12 corresponds to bit 1 of the digital output port (value 0x2).
	Pin 13 corresponds to bit 2 of the digital output port (value 0x4).
	Pin 14 corresponds to bit 3 of the digital output port (value 0x8).
	Pin 15 corresponds to bit 4 of the digital output port (value 0x10).
	Pin 16 corresponds to bit 5 of the digital output port (value 0x20).
	Pin 17 corresponds to bit 6 of the digital output port (value 0x40).
	Pin 18 corresponds to bit 7 of the digital output port (value 0x80).
Structure Element Name:	out_hi
Data Type:	uint8_t
Description:	Used by the <i>divider</i> and <i>one_shot</i> structures, specifies whether the output pulse for a rate generation or non-retriggerable one-shot operation is active high (1) or active low (0).
Structure Element Name:	retriggerable
Data Type:	uint8_t
Description:	Used by the <i>one-shot</i> structure, specifies whether or not to repeat the one-shot output pulse.
	Currently, only a value of 0 is supported.
	When <i>retriggerable</i> is 0, the counter generates a single pulse output signal whenever it detects the active gate signal (after the pulse period from the previous output pulse expires). Any gate signals that occur while the pulse is being output are not detected by the module. The module continues to output pulses until you stop the operation with ioctl - IOCTL_STOP_SUBSYS , described on page 139. You can use this mode to clean up a poor clock input signal by changing its pulse width, and then outputting it.
Returns	0 = Success; $< 0 =$ Failure
Notes	The value of the counter clock output signal (assigned to one of the general-purpose output signals) can be overwritten by writing to the digital output subsystem. Refer to page 53 for more information.
	If you assigned a general-purpose input signal as a counter clock or gate input (or external trigger), you can read the value of the signal as you would any other digital input signal. Refer to page 52 for more information on reading digital input values.

Example The following example sets up the counter/timer channel to generate a rate on the C/T clock output signal. In this example, the period of the output signal is 2 clock input signals of the internal 48 MHz C/T clock to generate a 24 MHz output signal and the pulse width of 1 clock input signal to generate a duty cycle of 50%. The operation is enabled when a logic high external gate signal is detected on general-purpose input signal 7. General-purpose output signal 7 is used for the C/T clock output signal. The output signal is active high.

```
#define CT_DEV_FILE
                                   "//dev/dt7837-ctr-tmr"
          ct_file = open(CT_DEV_FILE, O_RDWR);
          dt78xx_ct_config_t ct_cfg;
          ct_cfg.mode = ct_mode_divider;
          ct_cfg.gate = ct_gate_ext_hi
          ct_cfg.ext_gate_din = 0x80;
          ct_cfg.ext_clk = 0;
          ct_cfg.divider.period = 2;
          ct_cfg.divider.pulse = 1;
          ct cfg.divider.out = 0x80;
          ct_cfg.divider.out_high = 1;
          if (ioctl(ct_file, IOCTL_CT_CFG_SET, &ct_cfg))
              {
                  perror("IOCTL_CT_CFG_SET");
                  goto finish;
              }
See Also
          IOCTL_CT_CFG_GET, described on page 78.
```

ioctl - IOCTL_GAIN_POT_GET

Description	Returns the value of a specific potentiometer used for calibrating the gain.	
Syntax	<pre>int ioctl(int fd, IOCTL_GAIN_POT_GET, struct dt78xx_cal_pot_t *pGainCalPot);</pre>	
Include File	DT78XX_IOCTL.H	
Arguments		
Name:	fd	
Data Type:	int	
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).	
Name:	pGainCalPot	
Data Type:	dt78xx_cal_pot_t structure	
Description:	A pointer to a structure that defines the configuration of the calibration potentiometer.	
	<i>dt78xx_cal_pot_t</i> is defined as follows:	
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg; }cal; }dt78xx_cal_pot_t;</pre>	
Structure Element Name:	pot	
Data Type:	int	
Description:	The number of the channel to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, this value is 0.	
Structure Element Name:	value	
Data Type:	uint32_t	
Description:	The value of the potentiometer. Values range from 0 to 255.	
Structure Element Name:	factory	
Data Type:	int	
Description:	Identifies the calibration type, where 1 is factory calibration and 0 is user calibration.	

Structure Element Name:	reg
Data Type:	uint32_t
Description:	Identifies the calibration register. Values range from 0 to 15.
	For analog input calibration, the DT7837 uses registers 0 to 3, described as follows:
	• Register 0 – Stores the calibration factors when using a gain of 1 and a sampling frequency less than 52.734 kHz.
	• Register 1 – Stores the calibration factors when using a gain of 10 and a sampling frequency less than 52.734 kHz.
	• Register 2 – Stores the calibration factors when using a gain of 1 and a sampling frequency equal to or greater than 52.734 kHz.
	• Register 3 – Stores the calibration factors when using a gain of 10 and a sampling frequency equal to or greater than 52.734 kHz.
	For the analog output subsystem, the DT7837 uses registers 0 to 2, described as follows:
	• Register 0 – Stores the calibration factors when using an output frequency greater than 54 kHz and less than or equal to 108 kHz.
	• Register 1 – Stores the calibration factors when using an output frequency greater than 30 kHz and less than or equal to 54 kHz.
	• Register 2 – Stores the calibration factors when using an output frequency greater than 108 kHz and less than or equal to 216 kHz.
Returns	0 = Success; $< 0 = $ Failure
Example	The following example returns the value of the user calibration value of register 1 of the gain potentiometer for analog input 0.
	<pre>#define AIN_DEV_FILE</pre>
	<pre>gain_pot.pot = 0; gain_pot.value = 0; gain_pot.factory = 0; gain_pot.reg = 1;</pre>
	if (ain_file, IOCTL_GAIN_POT_GET, &gain_pot) {
	<pre>fprintf(stdout, "*** ERROR reading calibration pot eeprom from driver\n"); return false; }</pre>
	*value = gain_pot.value; return true;
See Also	IOCTL_GAIN_POT_SET, described on page 94.

ioctl - IOCTL_GAIN_POT_SET

Description	Sets the value of a specific potentiometer used to calibrate the gain.	
Syntax	<pre>int ioctl(int fd, IOCTL_GAIN_POT_SET, struct dt78xx_cal_pot_t *pGainCalPot);</pre>	
Include File	DT78XX_IOCTL.H	
Arguments		
Name:	fd	
Data Type:	int	
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).	
Name:	pGainCalPot	
Data Type:	dt78xx_cal_pot_t structure	
Description:	A pointer to a structure that defines the configuration of the calibration potentiometer.	
	<i>dt78xx_cal_pot_t</i> is defined as follows:	
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg; }cal; }dt78xx_cal_pot_t;</pre>	
Structure Element Name:	pot	
Data Type:	int	
Description:	The number of the channel to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, the value is 0.	
Structure Element Name:	value	
Data Type:	uint32_t	
Description:	The value of the potentiometer. Values range from 0 to 255.	
Structure Element Name:	factory	
Data Type:	int	
Description:	Identifies the calibration type, where 1 is factory calibration and 0 is user calibration. Typically, you want to modify the user calibration value; therefore, a value of 0 is recommended.	

Structure Element Name:	reg
Data Type:	uint32_t
Description:	Identifies the calibration register. Values range from 0 to 15.
	For the analog input subsystem, the DT7837 uses registers 0 to 3, described as follows:
	• Register 0 – Stores the calibration factors when using a gain of 1 and a sampling frequency less than 52.734 kHz.
	• Register 1 – Stores the calibration factors when using a gain of 10 and a sampling frequency less than 52.734 kHz.
	• Register 2 – Stores the calibration factors when using a gain of 1 and a sampling frequency equal to or greater than 52.734 kHz.
	• Register 3 – Stores the calibration factors when using a gain of 10 and a sampling frequency equal to or greater than 52.734 kHz.
	For the analog output subsystem, the DT7837 uses registers 0 to 2, described as follows:
	• Register 0 – Stores the calibration factors when using an output frequency greater than 54 kHz and less than or equal to 108 kHz.
	• Register 1 – Stores the calibration factors when using an output frequency greater than 30 kHz and less than or equal to 54 kHz.
	• Register 2 – Stores the calibration factors when using an output frequency greater than 108 kHz and less than or equal to 216 kHz.
Returns	0 = Success; $< 0 = $ Failure
Notes	You can continue to adjust the gain calibration values by calling this command with different values. The gain calibration register is automatically written to the wiper value of the associated potentiometer.
	This command will block all other operations for at least 5 ms.
Example	The following example sets the user calibration value for calibration register 0 of the gain potentiometer associated with analog input channel 0. In this example, the value is set to 255.
	<pre>#define AIN_DEV_FILE "//dev/dt7837-ain" ain_file = open(AIN_DEV_FILE, O_RDWR); dt78xx_cal_pot_t gain_pot; gain_pot.pot = 0; gain_pot.value = 255; gain_pot.factory = 0; gain_pot.reg =0; if (ioctl(ain_file, IOCTL_GAIN_POT_SET, &gain_pot)) {</pre>
	<pre>fprintf(stdout, "*** ERROR writing calibration pot eeprom to driver\n"); return false;</pre>
	}

See Also IOCTL_GAIN_POT_GET, described on page 92.

$ioctI - IOCTL_GAIN_POT_WIPER_GET$

Description	Returns the value of the wiper for a channel in a specific gain potentiometer.
Syntax	<pre>int ioctl(int fd, IOCTL_GAIN_POT_WIPER_GET, struct dt78xx_cal_pot_t *pGainCalWiperPot);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).
Name:	pGainCalWiperPot
Data Type:	dt78xx_cal_pot_t structure
Description:	A pointer to a structure that defines the configuration of the calibration potentiometers.
	<i>dt78xx_cal_pot_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg; }cal; }dt78xx_cal_pot_t;</pre>
Structure Element Name:	pot
Data Type:	int
Description:	The number of the channel/potentiometer to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, the value is 0.
Structure Element Name:	value
Data Type:	uint32_t
Description:	The value of the potentiometer. Values range from 0 to 255.
Structure Element Name:	factory
Data Type:	int
Description:	Not used for this command.

```
Structure Element Name:
                        reg
           Data Type:
                        uint32_t
          Description:
                        Not used for this command.
             Returns
                        0 = Success; < 0 = Failure
             Example
                        The following example returns the wiper value of the gain potentiometer
                        for analog input channel 0:
                        #define AIN_DEV_FILE
                                                    "//dev/dt7837-ain"
                        ain_file = open(AIN_DEV_FILE, O_RDWR);
                        dt78xx_cal_pot_t gain_wiper_pot;
                        gain_wiper_pot.pot = 0;
                        gain_wiper_pot.value = 0;
                        if (ain_file, IOCTL_GAIN_POT_WIPER_GET,
                           &gain_wiper_pot)
                          {
                              fprintf(stdout, "*** ERROR reading calibration
                                 pot eeprom from driver\n");
                              return false;
                           }
                        *value = gain_wiper_pot.value;
                        return true;
             See Also
```

IOCTL_GAIN_POT_WIPER_SET, described on page 99.

98

$ioctl - IOCTL_GAIN_POT_WIPER_SET$

Description	Sets the value of the wiper for a channel in a specific gain potentiometer	
Syntax	<pre>int ioctl(int fd, IOCTL_GAIN_POT_WIPER_SET, struct dt78xx_cal_pot_t *pGainCalWiperPot);</pre>	
Include File	DT78XX_IOCTL.H	
Arguments		
Name:	fd	
Data Type:	int	
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).	
Name:	pGainCalWiperPot	
Data Type:	dt78xx_cal_pot_t structure	
Description:	A pointer to a structure that defines the configuration of the calibration potentiometer.	
	<i>dt78xx_cal_pot_t</i> is defined as follows:	
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg; }cal; }dt78xx_cal_pot_t;</pre>	
Structure Element Name:	pot	
Data Type:	int	
Description:	The number of the channel/potentiometer to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, the value is 0.	
Structure Element Name:	value	
Data Type:	uint32_t	
Description:	The value of the potentiometer. Values range from 0 to 255.	
Structure Element Name:	factory	
Data Type:	int	
Description:	Not used for this command.	
Structure Element Name:	reg	
Data Type:	uint32_t	
Description:	Not used for this command.	

```
Returns
           0 = Success; < 0 = Failure
  Notes
           When you are satisfied that the gain calibration value is correct, call this
           command to update the calibration register in non-volatile EEPROM.
Example
           The following example sets the wiper value for the gain potentiometer
           associated with analog input channel 0 to a value of 255:
           #define AIN_DEV_FILE
                                         "//dev/dt7837-ain"
           ain_file = open(AIN_DEV_FILE, O_RDWR);
           dt78xx_cal_pot_t gain_wiper_pot;
           gain_wiper_pot.pot = 0;
           gain_wiper_pot.value = 255;
           if (ioctl(ain_file, IOCTL_GAIN_POT_WIPER_SET,
               &gain_wiper_pot))
                {
                    perror("IOCTL_GAIN_POT_WIPER_SET");
                     goto finish;
                }
See Also
           IOCTL_GAIN_POT_WIPER_GET, described on page 97.
```

ioctl - IOCTL_LED_GET

Description	Returns the configuration of the user LEDs on the DT7837 module.
Syntax	<pre>int ioctl(int fd, IOCTL_LED_GET, struct dt78xx_led_t *pLED_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with any of the subsystems on the DT7837 (/dev/dt7837-ain, /dev/dt7837-aout, /dev/dt7837-din, /dev/dt7837-dout, /dev/dt7837-tach, /dev/dt7837-measure, /dev/dt7837-ctr-tmr, /dev/dt7837-stream-in, or /dev/dt7837-stream-out).
Name:	pLED_config
Data Type:	dt78xx_led_t structure
Description:	A pointer to a structure that defines the configuration of the user LEDs on the DT7837 module.
	<i>dt78xx_led_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { uint8_t mask; uint8_t state; } dt78xx_led_t;</pre>
Structure Element Name:	mask
Data Type:	unit_8_t
Description:	An 8-bit unsigned value used by the ioctl - IOCTL_LED_SET command, described on page 103, to specify which user LEDs to modify. It is not used by the ioctl - IOCTL_LED_GET command.

Charles Elses of News	
Structure Element Name:	state
Data Type:	uint8_t
Description:	An 8-bit unsigned value indicating the status of the eight user LEDs on the DT7837 module.
	Each LED corresponds to a bit of the 8-bit value, as follows:
	• LED 7 corresponds to bit 7 (value 0x80).
	• LED 6 corresponds to bit 6 (value 0x40).
	• LED 5 corresponds to bit 5 (value 0x20).
	• LED 4 corresponds to bit 4 (value 0x10).
	• LED 3 corresponds to bit 3 (value 0x8).
	• LED 2 corresponds to bit 2 (value 0x4).
	• LED 1 corresponds to bit 1 (value 0x2).
	• LED 0 corresponds to bit 0 (value 0x1).
	A value of 0 indicates that the LED is off, and a value of 1 indicates that the LED is on.
Returns	0 = Success; $< 0 =$ Failure
Notes	Refer to Figure 5 on page 59 for the location of the user LEDs on the DT7837 module.
Example	The following example returns the configuration of the user LEDs on the DT7837 module:
	<pre>#define AIN_DEV_FILE "//dev/dt7837-ain" ain_file = open(AIN_DEV_FILE, O_RDWR); dt78xx_led_t LED_cfg;</pre>
	<pre>if (ioctl(ain_file, IOCTL_LED_GET, &LED_cfg)) { perror("IOCTL_LED_GET"); goto finish; }</pre>
See Also	IOCTL_LED_SET, described on page 103.

ioctl - IOCTL_LED_SET

Description	Specifies the configuration of the user LEDs on the DT7837 module.
Syntax	<pre>int ioctl(int fd, IOCTL_LED_SET, struct dt78xx_led_t *pLED_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with any of the subsystems on the DT7837 (/dev/dt7837-ain, /dev/dt7837-aout, /dev/dt7837-din, /dev/dt7837-dout, /dev/dt7837-tach, /dev/dt7837-measure, /dev/dt7837-ctr-tmr, /dev/dt7837-stream-in, or /dev/dt7837-stream-out).
Name:	pLED_config
Data Type:	dt78xx_led_t structure
Description:	A pointer to a structure that defines the configuration of the user LEDs on the DT7837 module.
	<i>dt78xx_led_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { uint8_t mask; uint8_t state; } dt78xx_led_t;</pre>
Structure Element Name:	mask
Data Type:	unit_8_t
Description:	An 8-bit unsigned value that specifies which of the eight user LEDs on the DT7837 module to modify. Each LED corresponds to a bit of the 8-bit value, as follows:
	• LED 7 corresponds to bit 7 (value 0x80).
	• LED 6 corresponds to bit 6 (value 0x40).
	• LED 5 corresponds to bit 5 (value 0x20).
	• LED 4 corresponds to bit 4 (value 0x10).
	• LED 3 corresponds to bit 3 (value 0x8).
	• LED 2 corresponds to bit 2 (value 0x4).
	• LED 1 corresponds to bit 1 (value 0x2).
	• LED 0 corresponds to bit 0 (value 0x1).
	Specify a value of 0 for the bit to keep the state of the LED unchanged. Specify a value of 1 if you want to modify the state of the LED.

Structure Element Name:	state
Data Type:	uint8_t
Description:	An 8-bit unsigned value indicating the state of the eight user LEDs on the DT7837 module. Each LED corresponds to a bit of the 8-bit value, as follows:
	• LED 7 corresponds to bit 7 (value 0x80).
	• LED 6 corresponds to bit 6 (value 0x40).
	• LED 5 corresponds to bit 5 (value 0x20).
	• LED 4 corresponds to bit 4 (value 0x10).
	• LED 3 corresponds to bit 3 (value 0x8).
	• LED 2 corresponds to bit 2 (value 0x4).
	• LED 1 corresponds to bit 1 (value 0x2).
	• LED 0 corresponds to bit 0 (value 0x1).
	Specify a value of 0 if you want to turn the LED off. Specify a value of 1 if you want to turn the LED on.
Returns	0 = Success; $< 0 =$ Failure
Notes	Refer to Figure 5 on page 59 for the location of the user LEDs on the DT7837 module.
Example	The following example modifies the state of user LEDs 0 and 2 on the DT7837 module (for a value of $0x5$). LED 0 is turned off and LED 2 is turned on (for a value of $0x4$):
	<pre>#define AIN_DEV_FILE "//dev/dt7837-ain" ain_file = open(AIN_DEV_FILE, O_RDWR); dt78xx_led_t LED_cfg;</pre>
	<pre>led_cfg.mask = 0x5; led_cfg.state = 0x4;</pre>
	<pre>if (ioctl(ain_file, IOCTL_LED_SET, &LED_cfg)) { perror("IOCTL_LED_SET"); goto finish; }</pre>
See Also	IOCTL_LED_GET , described on page 101.

ioctl - IOCTL_MCTR_CFG_GET

Description	Returns the configuration of the measure counter.
Syntax	<pre>int ioctl(int fd, IOCTL_MCTR_CFG_GET, struct dt78xx_mctr_config_t *pMCTR_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the measure counter subsystem (/dev/dt7837-measure).
Name:	pMCTR_config
Data Type:	dt78xx_mctr_config_t structure
Description:	A pointer to a structure that defines the configuration of the measure counter.
	<i>dt78xx_mctr_config_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { mctr_sel_t start_sel; mctr_sel_t stop_sel; uint8_t stale_flag; }; }dt78xx_mctr_config_t;</pre>

```
Structure Element Name:
                       start sel
           Data Type:
                       mctr_sel_t
          Description:
                       This enumeration defines the signal and edge that starts the measure
                       counter. It is defined as follows:
                       typedef enum
                       {
                         mctr_sel_adc_complete = 0x00,
                         mctr_sel_tach_falling = 0x10,
                         mctr_sel_tach_rising = 0x11,
                         mctr_sel_din0_falling = 0x20,
                         mctr_sel_din0_rising = 0x21,
                         mctr_sel_din1_falling = 0x22,
                         mctr sel din1 rising = 0x23,
                         mctr sel din2 falling = 0x24,
                         mctr_sel_din2_rising = 0x25,
                         mctr sel din3 falling = 0x26,
                         mctr_sel_din3_rising = 0x27,
                         mctr_sel_din4_falling = 0x28,
                         mctr sel din4 rising = 0x29,
                         mctr_sel_din5_falling = 0x2a,
                         mctr_sel_din5_rising = 0x2b,
                         mctr_sel_din6_falling = 0x2c,
                         mctr_sel_din6_rising = 0x2d,
                         mctr_sel_din7_falling = 0x2e,
                         mctr_sel_din7_rising = 0x2f,
                       } mctr_sel_t;
```

- *mctr_sel_adc_complete* (value 0x00) The measure counter starts when the A/D conversion is complete.
- *mctr_sel_tach_falling* (value 0x10) The measure counter starts on the falling edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_tach_rising* (value 0x11) The measure counter starts on the rising edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_din0_falling* (value 0x20) The measure counter starts on the falling edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- mctr_sel_din0_rising (value 0x21) The measure counter starts on the rising edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- mctr_sel_din1_falling (value 0x22) The measure counter starts on the falling edge of general-purpose input signal 1 (pin 2 of the Digital connector).

- *mctr_sel_din1_rising* (value 0x23) The measure counter starts on the rising edge of general-purpose input signal 1 (pin 2 of the Digital connector).
 - *mctr_sel_din2_falling* (value 0x24) The measure counter starts on the falling edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din2_rising* (value 0x25) The measure counter starts on the rising edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din3_falling* (value 0x26) The measure counter starts on the falling edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din3_rising* (value 0x27) The measure counter starts on the rising edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din4_falling* (value 0x28) The measure counter starts on the falling edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din4_rising* (value 0x29) The measure counter starts on the rising edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din5_falling* (value 0x2a) The measure counter starts on the falling edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din5_rising* (value 0x2b) The measure counter starts on the rising edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din6_falling* (value 0x2c) The measure counter starts on the falling edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din6_rising* (value 0x2d) The measure counter starts on the rising edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din7_falling* (value 0x2e) The measure counter starts on the falling edge of general-purpose input signal 7 (pin 8 of the Digital connector).
 - *mctr_sel_din7_rising* (value 0x2f) The measure counter starts on the rising edge of general-purpose input signal 7 (pin 8 of the Digital connector).

```
Structure Element Name:
                       stop_sel
           Data Type:
                       mctr_sel_t
          Description:
                       This enumeration defines the signal and edge that stops the measure
                       counter. It is defined as follows:
                       typedef enum
                       {
                         mctr_sel_adc_complete = 0x00,
                         mctr_sel_tach_falling = 0x10,
                         mctr_sel_tach_rising = 0x11,
                         mctr_sel_din0_falling = 0x20,
                         mctr_sel_din0_rising = 0x21,
                         mctr_sel_din1_falling = 0x22,
                         mctr sel din1 rising = 0x23,
                         mctr sel din2 falling = 0x24,
                         mctr_sel_din2_rising = 0x25,
                         mctr sel din3 falling = 0x26,
                         mctr_sel_din3_rising = 0x27,
                         mctr_sel_din4_falling = 0x28,
                         mctr sel din4 rising = 0x29,
                         mctr_sel_din5_falling = 0x2a,
                         mctr_sel_din5_rising = 0x2b,
                         mctr_sel_din6_falling = 0x2c,
                         mctr_sel_din6_rising = 0x2d,
                         mctr_sel_din7_falling = 0x2e,
                         mctr_sel_din7_rising = 0x2f,
                       } mctr_sel_t;
```

- *mctr_sel_adc_complete* (value 0x00) The measure counter stops when the A/D conversion is complete.
- *mctr_sel_tach_falling* (value 0x10) The measure counter stops on the falling edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_tach_rising* (value 0x11) The measure counter stops on the rising edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_din0_falling* (value 0x20) The measure counter stops on the falling edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- *mctr_sel_din0_rising* (value 0x21) The measure counter stops on the rising edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- *mctr_sel_din1_falling* (value 0x22) The measure counter stops on the falling edge of general-purpose input signal 1 (pin 2 of the Digital connector).
- *mctr_sel_din1_rising* (value 0x23) The measure counter stops on the rising edge of general-purpose input signal 1 (pin 2 of the Digital connector).
 - *mctr_sel_din2_falling* (value 0x24) The measure counter stops on the falling edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din2_rising* (value 0x25) The measure counter stops on the rising edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din3_falling* (value 0x26) The measure counter stops on the falling edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din3_rising* (value 0x27) The measure counter stops on the rising edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din4_falling* (value 0x28) The measure counter stops on the falling edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din4_rising* (value 0x29) The measure counter stops on the rising edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din5_falling* (value 0x2a) The measure counter stops on the falling edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din5_rising* (value 0x2b) The measure counter stops on the rising edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din6_falling* (value 0x2c) The measure counter stops on the falling edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din6_rising* (value 0x2d) The measure counter stops on the rising edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din7_falling* (value 0x2e) The measure counter stops on the falling edge of general-purpose input signal 7 (pin 8 of the Digital connector).
 - *mctr_sel_din7_rising* (value 0x2f) The measure counter stops on the rising edge of general-purpose input signal 7 (pin 8 of the Digital connector).

Structure Element Name:	stale_flag
Data Type:	uint8_t
Description:	Indicates whether or not the data is new. This flag is used only when the start edge and the stop edge is set to use the same pin and edge (such as pin 0 - DIN 0 rising as the start edge and pin 0 -DIN rising as the stop edge).
	If the <i>stale_flag</i> is 1 (Used), the most significant bit (MSB) of the value is set to 0 to indicate new data; reading the value before the measurement is complete returns an MSB of 1.
	If <i>stale_flag</i> is 0 (Not Used), the MSB is always set to 0.
Returns	0 = Success; $< 0 =$ Failure
Notes	By default, the general-purpose input signals are configured as digital input signals.
	You read the value of the measure counter through the input stream by setting bit 10 of the channel mask using the ioctl - IOCTL_CHAN_MASK_SET command, described on page 75.
Example	The following example returns the configuration of the measure counter to start on the rising edge of the tachometer and stop on the falling edge of general-purpose input signal 5. In this example, the general-purpose input signal is configured as a digital input line. The stale flag is not used.
	<pre>#define MCTR_DEV_FILE "//dev/dt7837-measure" mctr_file = open(MCTR_DEV_FILE, O_RDWR); dt78xx_mctr_config_t mctr_cfg;</pre>
	<pre>if (ioctl(mctr_file, IOCTL_MCTR_CFG_GET, &mctr_cfg)) { perror("IOCTL_MCTR_CFG_GET"); goto finish; }</pre>
See Also	IOCTL_MCT_CFG_SET , described on page 111.

ioctl - IOCTL_MCTR_CFG_SET

Description	Specifies the configuration of the measure counter.	
Syntax	<pre>int ioctl(int fd, IOCTL_MCTR_CFG_SET, struct dt78xx_mctr_config_t *pMCTR_config);</pre>	
Include File	DT78XX_IOCTL.H	
Arguments		
Name:	fd	
Data Type:	int	
Description:	The file descriptor associated with the measure counter subsystem (/dev/dt7837-measure).	
Name:	pMCTR_config	
Data Type:	dt78xx_mctr_config_t structure	
Description:	A pointer to a structure that defines the configuration of the measure counter.	
	<i>dt78xx_mctr_config_t</i> is defined as follows:	
	<pre>typedef structattribute ((packed)) { mctr_sel_t start_sel; mctr_sel_t stop_sel; uint8_t stale_flag; }; }dt78xx_mctr_config_t;</pre>	

```
Structure Element Name:
                       start sel
           Data Type:
                       mctr_sel_t
          Description:
                       This enumeration defines the signal and edge that starts the measure
                       counter. It is defined as follows:
                       typedef enum
                       {
                         mctr_sel_adc_complete = 0x00,
                         mctr_sel_tach_falling = 0x10,
                         mctr_sel_tach_rising = 0x11,
                         mctr_sel_din0_falling = 0x20,
                         mctr_sel_din0_rising = 0x21,
                         mctr_sel_din1_falling = 0x22,
                         mctr sel din1 rising = 0x23,
                         mctr sel din2 falling = 0x24,
                         mctr_sel_din2_rising = 0x25,
                         mctr sel din3 falling = 0x26,
                         mctr_sel_din3_rising = 0x27,
                         mctr_sel_din4_falling = 0x28,
                         mctr sel din4 rising = 0x29,
                         mctr_sel_din5_falling = 0x2a,
                         mctr_sel_din5_rising = 0x2b,
                         mctr_sel_din6_falling = 0x2c,
                         mctr_sel_din6_rising = 0x2d,
                         mctr_sel_din7_falling = 0x2e,
                         mctr_sel_din7_rising = 0x2f,
                       } mctr_sel_t;
```

where,

- *mctr_sel_adc_complete* (value 0x00) The measure counter starts when the A/D conversion is complete.
- *mctr_sel_tach_falling* (value 0x10) The measure counter starts on the falling edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_tach_rising* (value 0x11) The measure counter starts on the rising edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_din0_falling* (value 0x20) The measure counter starts on the falling edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- mctr_sel_din0_rising (value 0x21) The measure counter starts on the rising edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- mctr_sel_din1_falling (value 0x22) The measure counter starts on the falling edge of general-purpose input signal 1 (pin 2 of the Digital connector).

- *mctr_sel_din1_rising* (value 0x23) The measure counter starts on the rising edge of general-purpose input signal 1 (pin 2 of the Digital connector).
 - *mctr_sel_din2_falling* (value 0x24) The measure counter starts on the falling edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din2_rising* (value 0x25) The measure counter starts on the rising edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din3_falling* (value 0x26) The measure counter starts on the falling edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din3_rising* (value 0x27) The measure counter starts on the rising edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din4_falling* (value 0x28) The measure counter starts on the falling edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din4_rising* (value 0x29) The measure counter starts on the rising edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din5_falling* (value 0x2a) The measure counter starts on the falling edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din5_rising* (value 0x2b) The measure counter starts on the rising edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din6_falling* (value 0x2c) The measure counter starts on the falling edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din6_rising* (value 0x2d) The measure counter starts on the rising edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din7_falling* (value 0x2e) The measure counter starts on the falling edge of general-purpose input signal 7 (pin 8 of the Digital connector).
 - *mctr_sel_din7_rising* (value 0x2f) The measure counter starts on the rising edge of general-purpose input signal 7 (pin 8 of the Digital connector).

```
Structure Element Name:
                       stop_sel
           Data Type:
                       mctr_sel_t
          Description:
                       This enumeration defines the signal and edge that stops the measure
                       counter. It is defined as follows:
                       typedef enum
                       {
                         mctr_sel_adc_complete = 0x00,
                         mctr_sel_tach_falling = 0x10,
                         mctr_sel_tach_rising = 0x11,
                         mctr_sel_din0_falling = 0x20,
                         mctr_sel_din0_rising = 0x21,
                         mctr_sel_din1_falling = 0x22,
                         mctr sel din1 rising = 0x23,
                         mctr sel din2 falling = 0x24,
                         mctr_sel_din2_rising = 0x25,
                         mctr sel din3 falling = 0x26,
                         mctr_sel_din3_rising = 0x27,
                         mctr_sel_din4_falling = 0x28,
                         mctr sel din4 rising = 0x29,
                         mctr_sel_din5_falling = 0x2a,
                         mctr_sel_din5_rising = 0x2b,
                         mctr_sel_din6_falling = 0x2c,
                         mctr_sel_din6_rising = 0x2d,
                         mctr_sel_din7_falling = 0x2e,
                         mctr_sel_din7_rising = 0x2f,
                       } mctr_sel_t;
```

where,

- *mctr_sel_adc_complete* (value 0x00) The measure counter stops when the A/D conversion is complete.
- *mctr_sel_tach_falling* (value 0x10) The measure counter stops on the falling edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_tach_rising* (value 0x11) The measure counter stops on the rising edge of the tachometer input signal (pin 23 of the Digital connector).
- *mctr_sel_din0_falling* (value 0x20) The measure counter stops on the falling edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- *mctr_sel_din0_rising* (value 0x21) The measure counter stops on the rising edge of general-purpose input signal 0 (pin 1 of the Digital connector).
- *mctr_sel_din1_falling* (value 0x22) The measure counter stops on the falling edge of general-purpose input signal 1 (pin 2 of the Digital connector).

- *mctr_sel_din1_rising* (value 0x23) The measure counter stops on the rising edge of general-purpose input signal 1 (pin 2 of the Digital connector).
 - *mctr_sel_din2_falling* (value 0x24) The measure counter stops on the falling edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din2_rising* (value 0x25) The measure counter stops on the rising edge of general-purpose input signal 2 (pin 3 of the Digital connector).
 - *mctr_sel_din3_falling* (value 0x26) The measure counter stops on the falling edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din3_rising* (value 0x27) The measure counter stops on the rising edge of general-purpose input signal 3 (pin 4 of the Digital connector).
 - *mctr_sel_din4_falling* (value 0x28) The measure counter stops on the falling edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din4_rising* (value 0x29) The measure counter stops on the rising edge of general-purpose input signal 4 (pin 5 of the Digital connector).
 - *mctr_sel_din5_falling* (value 0x2a) The measure counter stops on the falling edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din5_rising* (value 0x2b) The measure counter stops on the rising edge of general-purpose input signal 5 (pin 6 of the Digital connector).
 - *mctr_sel_din6_falling* (value 0x2c) The measure counter stops on the falling edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din6_rising* (value 0x2d) The measure counter stops on the rising edge of general-purpose input signal 6 (pin 7 of the Digital connector).
 - *mctr_sel_din7_falling* (value 0x2e) The measure counter stops on the falling edge of general-purpose input signal 7 (pin 8 of the Digital connector).
 - *mctr_sel_din7_rising* (value 0x2f) The measure counter stops on the rising edge of general-purpose input signal 7 (pin 8 of the Digital connector).

Structure Element Name:	stale_flag
Data Type:	uint8_t
Description:	Indicates whether or not the data is new. This flag is used only when the start edge and the stop edge is set to use the same pin and edge (such as pin 0 - DIN 0 rising as the start edge and pin 0 -DIN rising as the stop edge).
	If the <i>stale_flag</i> is 1 (Used), the most significant bit (MSB) of the value is set to 0 to indicate new data; reading the value before the measurement is complete returns an MSB of 1.
	If <i>stale_flag</i> is 0 (Not Used), the MSB is always set to 0.
Returns	0 = Success; $< 0 = $ Failure
Notes	By default, the general-purpose input signals are configured as digital input signals.
	You read the value of the measure counter through the input stream by setting bit 10 of the channel mask using the ioctl - IOCTL_CHAN_MASK_SET command, described on page 75
Example	The following example sets up the measure counter to start on the rising edge of the tachometer and stop on the falling edge of general-purpose input signal 5. In this example, the general-purpose input signal is configured as a digital input line. The stale flag is not used.
	<pre>#define MCTR_DEV_FILE</pre>
	<pre>mctr_cfg.start_sel = mctr_sel_tach_rising; mctr_cfg.stop_sel = mctr_sel_din5_falling; mctr_cfg.stale_flag = o;</pre>
	<pre>if (ioctl(mctr_file, IOCTL_MCTR_CFG_SET, &mctr_cfg)) { perror("IOCTL_MCTR_CFG_SET"); goto finish; }</pre>
See Also	IOCTL_MCT_CFG_GET, described on page 105.

ioctl - IOCTL_OFFSET_POT_GET

Description	Returns the value of a specific potentiometer used to calibrate the offset value associated with a specific channel.
Syntax	<pre>int ioctl(int fd, IOCTL_OFFSET_POT_GET, struct dt78xx_cal_pot_t *pOffetCalPot);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).
Name:	pOffsetCalPot
Data Type:	dt78xx_cal_pot_t structure
Description:	A pointer to a structure that defines the configuration of the calibration potentiometer.
	<i>dt78xx_cal_pot_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg; }cal; }dt78xx_cal_pot_t;</pre>
Structure Element Name:	pot
Data Type:	int
Description:	The number of the channel/potentiometer to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, the value is 0.
Structure Element Name:	value
Data Type:	uint32_t
Description:	The value of the potentiometer. Values range from 0 to 255.
Structure Element Name:	factory
Data Type:	int
Description:	Identifies the calibration type, where 1 is factory calibration and 0 is user calibration.

Structure Element Name:	reg
Data Type:	uint32_t
Description:	Identifies the calibration register. Values range from 0 to 15.
	For the analog input subsystem, the DT7837 uses registers 0 to 3, described as follows:
	• Register 0 – Stores the calibration factors when using a gain of 1 and a sampling frequency less than 52.734 kHz.
	• Register 1 – Stores the calibration factors when using a gain of 10 and a sampling frequency less than 52.734 kHz.
	• Register 2 – Stores the calibration factors when using a gain of 1 and a sampling frequency equal to or greater than 52.734 kHz.
	• Register 3 – Stores the calibration factors when using a gain of 10 and a sampling frequency equal to or greater than 52.734 kHz.
	For the analog output subsystem, the DT7837 uses registers 0 to 2, described as follows:
	• Register 0 – Stores the calibration factors when using an output frequency greater than 54 kHz and less than or equal to 108 kHz.
	• Register 1 – Stores the calibration factors when using an output frequency greater than 30 kHz and less than or equal to 54 kHz.
	• Register 2 – Stores the calibration factors when using an output frequency greater than 108 kHz and less than or equal to 216 kHz.
Returns	0 = Success; $< 0 = $ Failure
Example	The following example returns the user calibration value of register 1 of the offset calibration potentiometer associated with analog input channel 0.
	<pre>#define AIN_DEV_FILE "//dev/dt7837-ain" ain_file = open(AIN_DEV_FILE, O_RDWR); dt78xx_cal_pot_t offset_pot;</pre>
	<pre>offset_pot.pot = 0; offset_pot.value = 0;</pre>
	offset pot.reg = 1;
	if (ain_file, IOCTL_OFFSET_POT_GET, &offset_pot)
	<pre>{ fprintf(stdout, "*** ERROR reading calibration pot eeprom from driver\n"); seture false</pre>
	}
	<pre>^value = offset_pot.value; return true;</pre>
See Also	IOCTL_OFFSET_POT_SET, described on page 119.

ioctl - IOCTL_OFFSET_POT_SET

Description	Sets the value of a specific potentiometer used to calibrate the offset value associated with a specific channel.
Syntax	<pre>int ioctl(int fd, IOCTL_OFFSET_POT_SET, struct dt78xx_cal_pot_t *pOffetCalPot);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).
Name:	pOffsetCalPot
Data Type:	dt78xx_cal_pot_t structure
Description:	A pointer to a structure that defines the configuration of the calibration potentiometer.
	<i>dt78xx_cal_pot_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg; }cal; }dt78xx_cal_pot_t;</pre>
Structure Element Name:	pot
Data Type:	int
Description:	The number of the channel/potentiometer to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, the value is 0.
Structure Element Name:	value
Data Type:	uint32_t
Description:	The value of the potentiometer. Values range from to 0 to 255.
Structure Element Name:	factory
Data Type:	int
Description:	Identifies the calibration type, where 1 is factory calibration and 0 is user calibration.

Structure Element Name:	reg
Data Type:	uint32_t
Description:	Identifies the calibration register. Values range from 0 to 15.
	For the analog input subsystem, the DT7837 uses registers 0 to 3, described as follows:
	 Register 0 – Stores the calibration factors when using a gain of 1 and a sampling frequency less than 52.734 kHz.
	• Register 1 – Stores the calibration factors when using a gain of 10 and a sampling frequency less than 52.734 kHz.
	• Register 2 – Stores the calibration factors when using a gain of 1 and a sampling frequency equal to or greater than 52.734 kHz.
	• Register 3 – Stores the calibration factors when using a gain of 10 and a sampling frequency equal to or greater than 52.734 kHz.
	For the analog output subsystem, the DT7837 uses registers 0 to 2, described as follows:
	• Register 0 – Stores the calibration factors when using an output frequency greater than 54 kHz and less than or equal to 108 kHz.
	• Register 1 – Stores the calibration factors when using an output frequency greater than 30 kHz and less than or equal to 54 kHz.
	• Register 2 – Stores the calibration factors when using an output frequency greater than 108 kHz and less than or equal to 216 kHz.
Returns	0 = Success; $< 0 = $ Failure
Notes	You can continue to adjust the offset calibration values by calling this command with different values. The offset calibration register is automatically written to the wiper of the associated potentiometer.
Example	The following example sets the user calibration value for calibration register 0 of the offset potentiometer associated with analog input channel 0. In this example, the value is set to 254.
	<pre>#define AIN_DEV_FILE "//dev/dt7837-ain" ain_file = open(AIN_DEV_FILE, O_RDWR); dt78xx_cal_pot_t offset_pot; offset_pot.pot = 0; offset_pot.value = 254; offset_pot.factory = 0; offset_pot.reg = 0;</pre>
	<pre>if (ioctl(ain_file, IOCTL_OFFSET_POT_SET, &offset_pot)) { fprintf(stdout, "*** ERROR writing calibration pot eeprom \n"); return false; }</pre>
See Also	IOCTL OFFSET POT CET described on page 117

ee Also **IOCTL_OFFSET_POT_GET**, described on page 117.

$ioctl - IOCTL_OFFSET_POT_WIPER_GET$

Description	Returns the value of the wiper for a channel in a specific offset potentiometer.
Syntax	<pre>int ioctl(int fd, IOCTL_OFFSET_POT_WIPER_GET, struct dt78xx_cal_pot_t *pOffsetCalWiperPot);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).
Name:	pOffsetCalWiperPot
Data Type:	dt78xx_cal_pot_t structure
Description:	A pointer to a structure that defines the configuration of the calibration potentiometers.
	<i>dt78xx_cal_pot_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg; }cal; }dt78xx_cal_pot_t;</pre>
Structure Element Name:	pot
Data Type:	int
Description:	The number of the channel/potentiometer to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, the value is 0.
Structure Element Name:	value
Data Type:	uint32_t
Description:	The value of the potentiometer. Values range from 0 to 255.
Structure Element Name:	factory
Data Type:	int
Description:	Not used for this command.

Structure Element Name: reg Data Type: uint32_t Description: Not used for this command. Returns 0 =Success; < 0 =Failure Example The following example returns the wiper value of the offset potentiometer for analog input channel 0: #define AIN_DEV_FILE "//dev/dt7837-ain" ain_file = open(AIN_DEV_FILE, O_RDWR); dt78xx_cal_pot_t offset_wiper_pot; offset_wiper_pot.pot = 0; offset_wiper_pot.value = 0; if (ain_file, IOCTL_OFFSET_POT_WIPER_GET, &offset_wiper_pot) { fprintf(stdout, "*** ERROR reading calibration pot eeprom from driver\n"); return false; } *value = offset_wiper_pot.value; return true; See Also **IOCTL_OFFSET_POT_WIPER_SET**, described on page 123.

$ioctl - IOCTL_OFFSET_POT_WIPER_SET$

Description	Sets the value of the wiper for a channel in a specific offset potentiometer.
Syntax	<pre>int ioctl(int fd, IOCTL_OFFSET_POT_WIPER_SET, struct dt78xx_cal_pot_t *pOffsetCalWiperPot);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the analog input subsystem (/dev/dt7837-ain) or analog output subsystem (/dev/dt7837-aout).
Name:	pOffsetCalWiperPot
Data Type:	dt78xx_cal_pot_t structure
Description:	A pointer to a structure that defines the configuration of the calibration potentiometer.
	<i>dt78xx_cal_pot_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { int pot; uint32_t value; struct { int factory; uint32_t reg;</pre>
	<pre>}cal; }dt78xx_cal_pot_t;</pre>
Structure Element Name:	pot
Data Type:	int
Description:	The number of the channel/potentiometer to calibrate. For the analog input subsystem, values range from 0 to 3. For the analog output subsystem, the value is 0.
Structure Element Name:	value
Data Type:	uint32_t
Description:	The value of the potentiometer. Values range from 0 to 255.
Structure Element Name:	factory
Data Type:	int
Description:	Not used for this command.
Structure Element Name:	reg
Data Type:	uint32_t
Description:	Not used for this command.

See Also

```
Returns
           0 = Success; < 0 = Failure
  Notes
           When you are satisfied that the offset calibration value is correct, call this
           command to update the calibration register in non-volatile EEPROM.
Example
           The following example sets the wiper for the offset potentiometer
           associated with analog input channel 0 to a value of 254:
           #define AIN_DEV_FILE
                                         "//dev/dt7837-ain"
           ain_file = open(AIN_DEV_FILE, O_RDWR);
           dt78xx_cal_pot_t offset_wiper_pot;
           offset_wiper_pot.pot = 0;
           offset_wiper_pot.value = 254;
           if (ioctl(ain_file, IOCTL_OFFSET_POT_WIPER_SET,
               &offset_wiper_pot))
              {
                 fprintf(stdout, "*** ERROR writing calibration
                    pot eeprom \n");
                 return false;
               }
```

IOCTL_OFFSET_POT_WIPER_GET, described on page 121.

124

$ioctl - IOCTL_SAMPLE_CLK_GET$

Description	Returns the configuration of the sample clock for the specified stream.
Syntax	<pre>int ioctl(int fd, IOCTL_SAMPLE_CLK_GET, dt78xx_clk_config_t *pClk_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or output stream (/dev/dt7837-stream-out).
Name:	pClk_config
Data Type:	dt78xx_clk_config_t
Description:	A pointer to a structure that defines the configuration of the clock.
	<i>dt78xx_clk_config_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { uint8_t ext_clk; uint8_t ext_clk_din; uint8_t unused[2]; float clk_freq; } dt78xx clk config t;</pre>
Structure Element Name:	ext clk
Data Type:	uint8_t
Description:	For the DT7837, this element is ignored as only an internal clock source is supported.
Structure Element Name:	ext_clk_din
Data Type:	uint8_t
Description:	For the DT7837 this element is ignored as only an internal clock is supported.
Structure Element Name:	unused[2]
Data Type:	uint8_t
Description:	Reserved for future use.
Structure Element Name:	clk_freq
Data Type:	float
Description:	Contains the actual sampling frequency that was set, in Hertz. For the input stream, values range from 195.3125 Hz to 105.469 kHz. For the output stream, values range from 30 kHz to 216 kHz.
Returns	0 = Success; $< 0 =$ Failure

See Also IOCTL_SAMPLE_CLK_SET, described on page 127.

ioctl - IOCTL_SAMPLE_CLK_SET

Description	Specifies the configuration of the sample clock for the specified stream.
Syntax	<pre>int ioctl(int fd, IOCTL_SAMPLE_CLK_SET, dt78xx_clk_config_t *pClk_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or output stream (/dev/dt7837-stream-out).
Name:	pClk_config
Data Type:	dt78xx_clk_config_t
Description:	A pointer to a structure that defines the configuration of the clock.
	<i>dt78xx_clk_config_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { uint8_t ext_clk; uint8_t ext_clk_din; uint8_t unused[2];</pre>
	<pre>float clk_freq; } dt78xx_clk_config_t;</pre>
Structure Element Name:	ext_clk
Data Type:	uint8_t
Description:	For the DT7837, this element is ignored as only an internal clock is supported.
Structure Element Name:	ext_clk_din
Data Type:	uint8_t
Description:	For the DT7837, this element is ignored as only an internal clock is supported.
Structure Element Name:	unused[2]
Data Type:	uint8_t
Description:	Reserved for future use.
Structure Element Name:	clk_freq
Data Type:	float
Description:	Specifies the sampling frequency, in Hertz. For the input stream, values range from 195.3125 Hz to 105.469 kHz. For the output stream, values range from 30 kHz to 216 kHz.
Returns	0 = Success; $< 0 = $ Failure

```
Notes
            The DT7837 driver sets the frequency of the internal clock as close as
            possible to the value that you specify. However, the value that you specify
            may not be the actual value that is set. To return the actual sample clock
            frequency that was set, use the ioctl - IOCTL_SAMPLE_CLK_SET
            command.
Example
            The following example specifies a sampling frequency of 100 kHz for the
            internal clock used by the input stream, fd_instream:
            dt78xx_clk_config_t clk_cfg;
            clk_cfg.clk_freq = 100000.0f;
            if (ioctl(fd_instream, IOCTL_SAMPLE_CLK_SET, &clk_cfg))
                 {
                     perror("IOCTL_SAMPLE_CLK_SET");
                     goto finish;
                 }
See Also
           IOCTL_SAMPLE_CLK_GET, described on page 125.
```

ioctl - IOCTL_START_SUBSYS

Description	Starts the I/O operation on the DT7837 module.
	When the file descriptor is associated with an input stream or output stream, this command starts the operation when a software trigger is specified; it has no effect if any other trigger source is specified.
	When the file descriptor is associated with a counter/timer subsystem, this command starts the counter/timer operation. If a software gate type is specified, the operation starts immediately. If the gate type is an external gate, the operation starts when the specified gate signal is detected. See the ioctl - IOCTL_CT_CFG_SET command, described on page 85, for more information.
Syntax	<pre>int ioctl(int fd, IOCTL_START_SUBSYS, unint32_t *pSimultaneous);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the file for the input stream (/dev/dt7837-stream-in), output stream (/dev/dt7837-stream-out), or counter/timer subsystem (/dev/dt7837-ctr-tmr) of the DT7837.
Name:	*pSimultaneous
Data Type:	unint32_t
Description:	A pointer to a variable that specifies whether or not to start the operations on the input and output streams simultaneously. By default, this value is NULL and operations on the input and output streams are not started simultaneously.
	If the file descriptor for the input stream is specified, a non-zero value starts operations on the input stream and the output stream simultaneously. If the file descriptor for the output stream is specified, a non-zero value starts the output stream and the input stream simultaneously.
	If the file descriptor for the counter/timer subsystem is specified, this value is NULL.
Returns	0 = Success; $< 0 =$ Failure
Notes	Once an operation is started, you can stop the operation using the ioctl - IOCL_STOP_SUBSYS command, described on page 139.

```
Example
           The following example configures the input stream, fd_instream, for a
           software trigger, arms the input stream, and then starts acquisition on the
           input stream.
           dt78xx_trig_config_t trig_cfg;
               trig_cfg.src = trig_src_sw;
               if (ioctl(fd_stream, IOCTL_START_TRIG_CFG_SET,
                   &trig_cfg))
                   {
                     perror("IOCTL_START_TRIG_CFG_SET");
                     goto finish;
                   }
           ret = ioctl(fd_instream, IOCTL_ARM_SUBSYS, 0);
           uint32_t simultaneous = 0;
           ret = ioctl(fd_instream, IOCTL_START_SUBSYS,
              &simultaneous);
See Also
           ioctl - IOCTL_STOP_SUBSYS, described on page 139.
```

ioctl - IOCTL_START_TRIG_CNF_GET

Description	Returns the configuration of the start trigger used by the specified stream.
Syntax	<pre>int ioctl(int fd, IOCTL_START_TRIG_CFG_GET, struct dt78xx_trig_config_t *pStartTrig_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or output stream (/dev/dt7837-stream-out).
Name:	pStartTrig_config
Data Type:	dt78xx_trig_config_t structure
Description:	A pointer to a structure that defines the configuration of the start trigger.
	<i>dt78xx_trig_config_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { trig_src_t src; union { structattribute ((packed)) { uint16_t edge_rising; uint16_t din; }ext; structattribute ((packed)) { uint16_t edge_rising; uint16_t ain; int32_t level; }threshold:</pre>
	<pre>} src_cig; } dt78xx_trig_config_t;</pre>

```
Structure Element Name:
                              src
              Data Type:
                              trig_src_t
             Description:
                              This enumeration defines the source of the start trigger. It is defined as
                              follows:
                              typedef enum
                              {
                                trig_src_sw,
                                trig_src_ext,
                                 trig_src_threshold
                              }trig_src_t;
                              where.
                              • trig_src_sw specifies a software trigger as the source of the start trigger.
                                 When this trigger source is specified, the operation starts when the ioctl
                                 - IOCTL_START_SUBSYS command, described on page 129, is
                                 executed.

    trig_src_ext specifies the external, digital (TTL) trigger as the source of

                                 the start trigger. The start trigger occurs when the device detects a
                                 transition on the active edge of the external digital (TTL) trigger signal.
                                 If you specify this trigger source, the values for the elements of the ext
                                 structure of dt78xx_trig_config_t, described below, are required.
                              • trig_src_threshold specifies the threshold trigger as the source of the start
                                 trigger. The start trigger occurs when the signal attached to a specified
                                 analog input channel rises above or falls below a user-specified
                                 threshold value. If you specify this trigger source, the values for the
                                 elements of the threshold structure of dt78xx_trig_config_t, described on
                                 below, are required.
Structure Element Name:
                              edge_rising
              Data Type:
                              uint16 t
             Description:
                              Used by the ext and threshold structures, specifies which edge of the trigger
                              signal is the active edge, where 0 is falling edge and 1 is rising edge.
                              If this element is specified in the ext structure, the trigger signal refers the
                              external digital (TTL) trigger.
                              If this element is specified in the threshold structure, the trigger signal refers
                              the threshold trigger.
```

Structure Element Name:	din
Data Type:	uint16_t
Description:	Used by the <i>ext</i> structure, specifies which general-purpose input signal (pins 1 to 8) on the Digital connector to use for the external trigger input signal.
	Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
	Pin 2 corresponds to bit 1 of the digital input port (value 0x2).
	Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
	Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
	Pin 5 corresponds to bit 4 of the digital input port (value 0x10).
	Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
	Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
	Pin 8 corresponds to bit 7 of the digital input port (value 0x80).
Structure Element Name:	ain
Data Type:	uint16_t
Description:	Used by the <i>threshold</i> structure, specifies which analog input channel to use for the threshold trigger, where 0 represents analog input channel 0, 1 represents analog input channel 1, 2 represents analog input channel 2, and 3 represents analog input channel 3.
Structure Element Name:	level
Data Type:	int32_t
Description:	Specifies the level at which the threshold trigger occurs. Values depend on the resolution of the module. Since the DT7837 module has a 24-bit A/D converter and uses two's complement data encoding, the positive full-scale range (+10 V or +1 V depending on the gain) is 0x007fffff, one code above 0 is 0x0, one code below 0 is 0xff800000, and the negative full-scale range (-10 V or -1 V depending on the gain) is 0xffffffff.
Returns	0 = Success; $< 0 = $ Failure
Notes	The threshold trigger is supported for the input stream only.
	The DT7837 driver sets the threshold level as close as possible to the value that you specify. However, the value that you specify may not be the actual value that is set. This command returns the actual threshold level that was set using the ioctl - IOCTL_START_TRIG_CNF_SET command.

```
Example The following example returns the configuration of the start trigger for the
input stream, fd_instream:
    dt78xx_trig_config_t StartTrig_cfg;
    if (ioctl(fd_instream, IOCTL_START_TRIG_CFG_GET,
        &StartTrig_cfg))
        {
            perror("IOCTL_START_TRIG_CFG_GET");
            goto finish;
        }
See Also IOCTL_START_TRIG_CFG_SET, described on page 135.
```

ioctl - IOCTL_START_TRIG_CNF_SET

Description	Specifies the configuration of the start trigger used by the specified stream.
Syntax	<pre>int ioctl(int fd, IOCTL_START_TRIG_CFG_SET, struct dt78xx_trig_config_t *pStartTrig_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in) or output stream (/dev/dt7837-stream-out).
Name:	pStartTrig_config
Data Type:	dt78xx_trig_config_t structure
Description:	A pointer to a structure that defines the configuration of the start trigger.
	<i>dt78xx_trig_config_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { trig_src_t src; union { structattribute ((packed)) { uint16_t edge_rising; uint16_t din; }ext;</pre>
	<pre>structattribute ((packed)) { uint16_t edge_rising; uint16_t ain; int32_t level; }threshold; } src_cfg; } dt78xx_trig_config_t;</pre>

```
Structure Element Name:
                              src
              Data Type:
                              trig_src_t
             Description:
                              This enumeration defines the source of the start trigger. It is defined as
                              follows:
                              typedef enum
                              {
                                trig_src_sw,
                                trig_src_ext,
                                trig_src_threshold
                              }trig_src_t;
                              where.
                              • trig_src_sw specifies a software trigger as the source of the start trigger.
                                 When this trigger source is specified, the operation starts when the ioctl
                                 - IOCTL_START_SUBSYS command, described on page 129, is
                                 executed.
                                trig_src_ext specifies the external, digital (TTL) trigger as the source of
                              ٠
                                 the start trigger. The start trigger occurs when the device detects a
                                 transition on the active edge of the external digital (TTL) trigger signal.
                                 If you specify this trigger source, the values for the elements of the ext
                                 structure of dt78xx_trig_config_t, described below, are required.
                              • trig_src_threshold specifies the threshold trigger as the source of the start
                                 trigger. The start trigger occurs when the signal attached to a specified
                                 analog input channel rises above or falls below a user-specified
                                 threshold value. If you specify this trigger source, the values for the
                                 elements of the threshold structure of dt78xx_trig_config_t, described on
                                 below, are required.
Structure Element Name:
                              edge_rising
              Data Type:
                              uint16 t
             Description:
                              Used by the ext and threshold structures, specifies which edge of the trigger
                              signal is the active edge, where 0 is falling edge and 1 is rising edge.
                              If this element is specified in the ext structure, the trigger signal refers the
                              external digital (TTL) trigger.
                              If this element is specified in the threshold structure, the trigger signal refers
                              the threshold trigger.
```

Structure Element Name:	din
Data Type:	uint16_t
Description:	Used by the <i>ext</i> structure, specifies which general-purpose input signal (pins 1 to 8) on the Digital connector to use for the external trigger input signal.
	Pin 1 corresponds to bit 0 of the digital input port (value 0x1).
	Pin 2 corresponds to bit 1 of the digital input port (value 0x2).
	Pin 3 corresponds to bit 2 of the digital input port (value 0x4).
	Pin 4 corresponds to bit 3 of the digital input port (value 0x8).
	Pin 5 corresponds to bit 4 of the digital input port (value 0x10).
	Pin 6 corresponds to bit 5 of the digital input port (value 0x20).
	Pin 7 corresponds to bit 6 of the digital input port (value 0x40).
	Pin 8 corresponds to bit 7 of the digital input port (value 0x80).
Structure Element Name:	ain
Data Type:	uint16_t
Description:	Used by the <i>threshold</i> structure, specifies which analog input channel to use for the threshold trigger, where 0 represents analog input channel 0, 1 represents analog input channel 1, 2 represents analog input channel 2, and 3 represents analog input channel 3.
Structure Element Name:	level
Data Type:	uint32_t
Description:	Specifies the level at which the threshold trigger occurs. Values depend on the resolution of the module. Since the DT7837 module has a 24-bit A/D converter and uses two's complement data encoding, the positive full-scale range (+10 V or +1 V depending on the gain) is 0x007fffff, one code above 0 is 0x0, one code below 0 is 0xff800000, and the negative full-scale range (-10 V or -1 V depending on the gain) is 0xffffffff.
Returns	0 = Success; $< 0 = $ Failure
Notes	The threshold trigger is supported for the input stream only.
	The DT7837 driver sets the threshold level as close as possible to the value that you specify. However, the value that you specify may not be the actual value that is set. To return the actual threshold level that was set, use the ioctl - IOCTL_START_TRIG_CNF_GET command.

Example The following example specifies the configuration of the start trigger for the input stream, *fd_instream*. In this example, a rising-edge threshold trigger on analog input channel 0 is used with a threshold level of approximately +0.5 V (a two's complement value of 0x0066667): dt78xx_trig_config_t StartTrig_cfg; StartTrig_cfg.src = trig_src_threshold; StartTrig_cfg.src_cfg.threshold.edge_rising = 1; StartTrig_cfg.src_cfg.threshold.ain = 0; StartTrig_cfg.src_cfg.threshold.level = 0x0066667; if (ioctl(fd_instream, IOCTL_START_TRIG_CFG_SET, &StartTrig_cfg)) { perror("IOCTL_START_TRIG_CFG_SET"); goto finish; }

See Also IOCTL_START_TRIG_CFG_GET, described on page 131.

ioctl - IOCTL_STOP_SUBSYS

Description	Stops the I/O operation on the DT7837 module immediately.
Syntax	int ioctl(int fd, IOCTL_STOP_SUBSYS, int unused);
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the input stream (/dev/dt7837-stream-in), output stream (/dev/dt7837-stream-out), or counter/timer subsystem (/dev/dt7837-ctr-tmr) of the DT7837.
Name:	unused
Data Type:	int
Description:	Not used; reserved for future use.
Returns	0 = Success; $< 0 = $ Failure
Notes	When the file descriptor is associated with a counter/timer subsystem, this command stops an operation that was previously started with the ioctl - IOCTL_START_SUBSYS command, described on page 129.
	When the file descriptor is associated with an input stream or output stream, this command stops an operation that was previously started with the ioctl - IOCTL_START_SUBSYS command (a software trigger was specified) or was started when the specified trigger condition was detected. Refer to page 129 for more information on ioctl - IOCTL_START_SUBSYS . This command stops the DMA engine immediately and no further data is collected. Asynchronous I/O control blocks that were submitted using io_submit are still in the AIO queue and will not be completed. To cancel these control blocks, use io_cancel .
	If you want to restart the operation for the input stream or output stream, you must rearm the stream using ioctl - IOCTL_ARM_SUBSYS , described on page 71, and, if a software trigger is specified, restart the operation using ioctl - IOCTL_START_SUBSYS , described on page 129.
	Before terminating the application, use free to free any events and I/O control blocks and io_destroy to destroy the asynchronous I/O context used by the stream.
Example	The following example stops acquisition on the input stream, <i>fd_instream</i> :
	ioctl(fd_instream, IOCTL_STOP_SUBSYS, 0);

ioctl - IOCTL_TACH_CFG_GET

Description	Returns the configuration of the tachometer subsystem.
Syntax	<pre>int ioctl(int fd, IOCTL_TACH_CFG_GET, struct dt78xx_tach_config_t *pTach_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the tachometer subsystem (/dev/dt7837-tach).
Name:	pTach_config
Data Type:	dt78xx_tach_config_t structure
Description:	A pointer to a structure that defines the configuration of the tachometer subsystem.
	<i>dt78xx_tach_config_t</i> is defined as follows:
	<pre>typedef structattribute ((packed)) { uint8_t edge_rising; uint8_t stale_flag; }; }dt78xx_tach_config_t;</pre>
Structure Element Name:	edge_rising
Data Type:	uint8_t
Description:	Specifies whether the rising edge or falling edge of the tachometer is used for the measurement. The number of counts between two consecutive edges of the tachometer input signal is used as the measurement.
	If <i>edge_rising</i> is 0, the falling edge of the tachometer input signal is used for the measurement.
	If <i>edge_rising</i> is 1, the rising edge of the tachometer input signal is used for the measurement.
Structure Element Name:	stale_flag
Data Type:	uint8_t
Description:	Indicates whether or not the data is new.
	If the <i>stale_flag</i> is 1 (Used), the most significant bit (MSB) of the value is set to 0 to indicate new data; reading the value before the measurement is complete returns an MSB of 1.
	If <i>stale_flag</i> is 0 (Not Used), the MSB is always set to 0.
Returns	0 = Success; < 0 = Failure

Notes	You read the tachometer measurement through the input stream by setting bit 8 of the channel mask using the ioctl - IOCTL_CHAN_MASK_SET command, described on page 75.
Example	The following example returns the configuration of the tachometer subsystem, <i>tach_file</i> :
	if (ioctl(tach_file, IOCTL_TACH_CFG_GET, &tach_cfg)) {
	<pre>perror("IOCTL_TACH_CFG_GET");</pre>
	goto finish;
	}
See Also	IOCTL_TACH_CFG_SET, described on page 142.

ioctl - IOCTL_TACH_CFG_SET

Description	Specifies the configuration of the tachometer subsystem.
Syntax	<pre>int ioctl(int fd, IOCTL_TACH_CFG_SET, struct dt78xx_tach_config_t *pTACH_config);</pre>
Include File	DT78XX_IOCTL.H
Arguments	
Name:	fd
Data Type:	int
Description:	The file descriptor associated with the tachometer subsystem (/dev/dt7837-tach).
Name:	pTach_config
Data Type:	dt78xx_tach_config_t structure
Description:	A pointer to a structure that defines the configuration of the tachometer subsystem.
	<i>dt78xx_tach_config_t</i> is defined as follows:
	typedef structattribute ((packed))
	<pre>uint8_t edge_rising; uint8_t stale_flag; };</pre>
	}dt78xx_tach_config_t;
Structure Element Name:	edge_rising
Data Type:	uint8_t
Description:	Specifies whether the rising edge or falling edge of the tachometer is used for the measurement. The number of counts between two consecutive edges of the tachometer input signal is used as the measurement.
	If <i>edge_rising</i> is 0, the falling edge of the tachometer input signal is used for the measurement.
	If <i>edge_rising</i> is 1, the rising edge of the tachometer input signal is used for the measurement.
Structure Element Name:	stale_flag
Data Type:	uint8_t
Description:	Indicates whether or not the data is new.
	If the <i>stale_flag</i> is 1 (Used), the most significant bit (MSB) of the value is set to 0 to indicate new data; reading the value before the measurement is complete returns an MSB of 1.
	If <i>stale_flag</i> is 0 (Not Used), the MSB is always set to 0.
Returns	0 = Success; < 0 = Failure

- **Notes** You read the tachometer measurement through the input stream by setting bit 8 of the channel mask using the **ioctl IOCTL_CHAN_MASK_SET** command, described on page 75.
- **Example** The following example sets up the tachometer subsystem to measure between two consecutive rising edges of the tachometer signal. In this example, the stale flag is used.

```
#define TACH_DEV_FILE "//dev/dt7837-tach"
tach_file = open(TACH_DEV_FILE, O_RDWR);
dt78xx_tach_config_t tach_cfg;
tach_cfg.edge_rising = 1;
tach_cfg.stale_flag = 1;
if (ioctl(tach_file, IOCTL_TACH_CFG_SET, &tach_cfg))
{
    perror("IOCTL_TACH_CFG_SET");
    goto finish;
}
```

See Also IOCTL_TACH_CFG_GET, described on page 140.

open

Description	Given a pathname to a file that represents a subsystem, stream, or endpoint of the DT7837 module, opens the file and returns an associated file descriptor for use in subsequent calls.
Syntax	<pre>int open(const char *pathname, int flags);</pre>
Arguments	
Name:	*pathname
Data Type:	const char
Description:	A pointer to a filename that specifies the file to open. The following pathnames are supported by the DT7837 module:
	 /dev/dt7837-ain – Analog input subsystem
	 /dev/dt7837-aout – Analog output subsystem
	 /dev/dt7837-din – Digital input subsystem
	 /dev/dt7837-dout – Digital output subsystem
	• /dev/dt7837-tach – Tachometer subsystem
	• /dev/dt7837-measure – Measure counter subsystem
	• /dev/dt7837-ctr-tmr – Counter/timer subsystem
	 /dev/dt7837-stream-in – Input stream
	 /dev/dt7837-stream-out – Output stream
	 /dev/dt7837-ep1in – EP1 IN endpoint, address 0x81
	 /dev/dt7837-ep1out – EP1 OUT endpoint, address 0x01
	 /dev/dt7837-ep2in – EP2 IN endpoint, address 0x82
	 /dev/dt7837-ep2out – EP2 OUT endpoint, address 0x02
	 /dev/dt7837-ep3out – EP3 OUT endpoint, address 0x03
	 /dev/dt7837-ep4out – EP4 OUT endpoint, address 0x04
	 /dev/dt7837-ep5out – EP5 OUT endpoint, address 0x05
Name:	flags
Data Type:	int
Description:	The access mode of the file to open. The following access modes are supported:
	• O_RDONLY – Open the file for read only.
	• O_WRONLY – Open the file for write only.
	• O_RDWR – Open the file for read and write.
Returns	The new file descriptor for the file or -1 if an error occurred.
Notes	You must open the file corresponding to each subsystem, stream, or endpoint that you want to use before you can configure its parameters and perform an I/O read or write operation.
----------	--
	A file can be opened by only one reader/writer at a time. However, different files can be opened simultaneously. Once a file is closed, it can be re-opened.
	The file descriptor remains open until is closed with the close command, described on page 65.
Example	The following command opens the file associated with the input stream of the DT7837 for read only and returns a file descriptor that is associated with the input stream:
	#define DEV_STREAM_IN "/dev/dt7837-stream-in" fd_instream = open(DEV_STREAM_IN, O_RDONLY);
See Also	close , described on page 65.

Programming Flowcharts Using the File I/O Commands

Input Stream Asynchronous Read Operations	. 148
Analog Output Synchronous Write Operation	. 151
Output Stream Asynchronous Write Operations	. 152
Digital Input Synchronous Read Operation	. 154
Digital Output Synchronous Write Operation	. 155
Counter/Timer Operation	. 156
Calibration	. 157
User LED Modifications	. 158
Sending Data to a USB Host	. 159
Receiving Data from a USB Host	. 160

Input Stream Asynchronous Read Operations







Analog Output Synchronous Write Operation



Output Stream Asynchronous Write Operations



To start a continuous input operation and a continuous output operation simultaneously, specify a non-zero value for the *pSimultaneous* variable



Digital Input Synchronous Read Operation



Digital Output Synchronous Write Operation



Counter/Timer Operation



Calibration



User LED Modifications



Sending Data to a USB Host



Receiving Data from a USB Host





Product Support

Should you experience problems using the file I/O commands to program a DT7837 module, follow these steps:

- 1. Read all the appropriate sections of this manual and the DT7837 User's Manual.
- 2. Refer to the supplied example programs for clarification.
- **3.** Check that you have installed your hardware devices properly. For information, refer to the *DT7837 User's Manual*.
- **4.** Check that you have installed the software properly. For information, refer to the *DT7837 User's Manual*.

If you are still having problems, Data Translation's Technical Support Department is available to provide technical assistance. To request technical support, go to our web site at http://www.mccdaq.com and click on the Support link.

When requesting technical support, be prepared to provide the following information:

- Your product serial number
- The hardware/software product you need help on

If you are located outside the USA, contact your local distributor; see our web site (www.mccdaq.com) for the name and telephone number of your nearest distributor.

Index

A

acquisition status 28 AIO model 27, 34 aio_read 62 USB gadget 160 aio_write 61 USB gadget 159 aliasing 25 analog input calibration 54 analog input channels 24 analog input operations 24 configuring input channels 24 flowchart 148 opening the subsystem 24 analog output calibration 56 analog output operations 32, 152 flowchart 151 opening the subsystem 32 asynchronous I/O context 27, 34 asynchronous I/O requests 31, 38 asynchronous read operations 148 asynchronous write operations 152

B

blocks, I/O control 27, 28, 34, 35 buffer error reporting 29, 35 buffers input 28 output 35

С

calibration 54 analog input 54 analog output 56 flowchart 157 gain 56, 58 offset 55, 57 opening the subsystem 55, 57 channel mask 25, 33 clock sources counter/timer 45 input stream 25 output stream 33 close 23, 60, 65, 150 analog input 31, 38, 56, 58, 157 analog output 151 counter/timer 47, 156 digital input 52, 154 digital output 32, 53, 155 input stream 31, 38, 150 output stream 153 USB gadget 62, 159, 160 control blocks 27, 28, 34, 35 counter/timer operations 41 clock sources 45 closing the subsystem 47 configuring the subsystem 41 flowchart 156 flowchart for input stream operations 148 gate types 44 mode 41 opening the subsystem 41 polarity of the output signal 46 pulse output period 46 pulse width 46 reading the counter/timer 47 starting the operation 47 stopping the operation 47 coupling type 24 current source 24

D

debug pins on header J10 59 digital input operations 52 closing the subsystem 52 flowchart for a single read operation 154 flowchart for input stream operations 148 opening the subsystem 52 reading the value 52 digital output operations 53, 152 closing the subsystem 32, 53, 56, 58 flowchart 155 opening the subsystem 32, 53 updating the value 32, 53 digital trigger 26, 33 DT7837 device driver 10, 22 duty cycle 46

Ε

enabling buffer error reporting 29, 35 endpoint files 22 endpoints, USB 61 error reporting 29, 35 event counting mode 41 events 28, 35 external C/T clock source 45 external digital trigger 26, 33

F

files, I/O 22 flowcharts analog input operations 148 analog output operations 151, 152 calibration 157 counter/timer operations 156 digital input single read operations 154 digital output operations 155 input stream operations 148 output stream operations 152 receiving data from a USB host 160 sending data to a USB host 159 user LEDs 158 flush 62 USB gadget 159, 160 flushing data 62 frequency external C/T clock 45 sample clock 25, 33

G

gain 24 gain calibration 56, 58 gate type 44 group delay 25, 33

Η

header J10 59 help 8

|

I/O control blocks 27, 28, 34, 35 I/O files 22 I/O requests 27, 28, 34, 35 idle mode 44 input stream operations 24 arming the operation 28 cleaning up resources 31 configuring the channel mask 25 configuring the sample clock 25 configuring the start trigger 26 flowchart 148 opening the input stream 24 processing I/O requests 28, 35 starting the operation 28 stopping the operation 31 submitting I/O requests 27, 34 input type 24 internal C/T clock source 45 io_cancel 31, 38, 150, 153 io_destroy 31, 38, 150, 153 io_getevents 28, 35, 149, 153 io_setup 27, 34, 149, 152 io_submit 27, 34, 149, 152 ioctl - IOCTL_ACQ_STATUS_GET 28, 35, 66, 149, 152 ioctl - IOCTL_AIN_CFG_GET 24, 67 ioctl - IOCTL_AIN_CFG_SET 24, 69, 148 ioctl - IOCTL_ARM_SUBSYS 28, 34, 71, 149, 152 ioctl - IOCTL CHAN MASK GET 25, 33, 72 ioctl - IOCTL_CHAN_MASK_SET 25, 33, 75, 148, 152 ioctl - IOCTL CT CFG GET 41, 78 ioctl - IOCTL_CT_CFG_SET 41, 85, 148, 156 ioctl - IOCTL_GAIN_POT_GET 56, 58, 92 ioctl - IOCTL GAIN POT SET 56, 58, 94, 157 ioctl - IOCTL_GAIN_POT_WIPER_GET 56, 58, 97 ioctl - IOCTL_GAIN_POT_WIPER_SET 56, 58, 99, 157 ioctl - IOCTL_LED_GET 60, 101 ioctl - IOCTL_LED_SET 60, 103, 158 ioctl - IOCTL_MCTR_CFG_GET 51, 105 ioctl - IOCTL_MCTR_CFG_SET 48, 111, 148 ioctl - IOCTL OFFSET POT GET 55, 57, 117 ioctl - IOCTL_OFFSET_POT_SET 55, 57, 119, 157 ioctl-IOCTL_OFFSET_POT_WIPER_GET 55, 57, 121 ioctl - IOCTL_OFFSET_POT_WIPER_SET 55, 57, 123, 157 ioctl - IOCTL_SAMPLE_CLK_GET 25, 33, 125, 128 ioctl - IOCTL_SAMPLE_CLK_SET 25, 33, 127, 149, 152

ioctl - IOCTL_START_SUBSYS 28, 34, 129
 counter/timer 47, 156
 input stream 26, 33, 149
 output stream 152
ioctl - IOCTL_START_TRIG_CFG_GET 27, 34,
 131, 137
ioctl - IOCTL_START_TRIG_CFG_SET 26, 33,
 135, 149, 152
ioctl - IOCTL_STOP_SUBSYS 31, 37, 139
 counter/timer 47, 156
 input stream 149
 output stream 153
ioctl - IOCTL_TACH_CFG_GET 40, 140
ioctl - IOCTL_TACH_CFG_SET 39, 142

J

J10 header 59

L

LEDs, user 59 flowcharts 158

М

mask, channel 25, 33 measure counter operations 48 configuring the subsystem 48 flowchart for input stream operations 148 opening the subsystem 48

Ν

non-retriggerable one-shot mode 43 Nyquist Theorem 25

0

offset calibration 55, 57 **open** 23, 60, 144 analog input 24, 32, 55, 57, 148, 157, 158 analog output 151 counter/timer 41, 156, 159, 160 digital input 52, 154 digital output 32, 53, 155 input stream 24, 32, 148 measure counter 48, 148 output stream 152 tachometer 39, 148 USB gadget 61 output operation status 35 output stream operations 32 arming the operation 34 cleaning up resources 38 configuring the channel mask 33 configuring the sample clock 33 configuring the start trigger 33 flowchart 152 opening the output stream 32 starting the operation 34 stopping the operation 37 overrun errors 29

Ρ

polarity of the counter output signal 46 processing requests 27, 28, 34, 35 pulse output period 46 pulse width 46

R

rate generation mode 42 read counter/timer 47, 156 digital input 52, 154 USB gadget 62, 160 receiving data from a USB host 160 receiving data from the USB host 62 related documents 8 removing data 62 requests 27, 28, 34, 35

S

sample clock frequency 25, 33 sending data to a USB host 159 sending data to the USB host 61 service and support procedure 162 software trigger 26, 33 start trigger external digital (TTL) trigger 26, 33 software 26, 33 threshold trigger 26 status of acquisition 28 status of output operation 35 stream files 22 submitting requests 27, 34 subsystem files 22 synchronous read operation counter/timer 156 digital input 154 synchronous write operations 151, 155

T

tachometer operations 39 configuring the tachometer subsystem 39 flowchart for input stream operations 148 opening the subsystem 39 technical support 8, 162 threshold trigger 26 triggers external 26, 33 software 26, 33 threshold 26 troubleshooting checklist 162 troubleshooting procedure 162 TTL trigger 26, 33

U

underrun errors 35 USB endpoints 61 USB gadget closing the file 62 flowchart for receiving data from a USB host 160 flowchart for sending data to a USB host 159 opening the file 61 receiving data from the USB host 62 removing data 62 sending data to the USB host 61 USB gadget driver 10, 22 USB host 61, 62 user LEDs 59 flowchart 158

W

```
write
analog output 151
digital output 32, 53, 155
USB gadget 61, 159
```